

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P4763

Contract: VERI01

Lab Code: CHEM

Case No.: P4763

SAS No.: P4763

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164796BS							
Chromium	ug/L	200	197		98	80 - 120	P
Copper	ug/L	150	158		105	80 - 120	P
Nickel	ug/L	250	246		98	80 - 120	P
Zinc	ug/L	100	107		107	80 - 120	P