

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : sohil

Run Number: LB133337

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/07/2024	14:00
2	CAL2	1	Water	NA	NA	20.2	7.00	11/07/2024	14:02
3	CAL3	1	Water	NA	NA	20.3	10.02	11/07/2024	14:05
4	ICV	1	Water	NA	NA	20.1	7.01	11/07/2024	14:10
5	CCV1	1	Water	NA	NA	20.2	2.01	11/07/2024	14:15
6	P4759-01	1	Water	NA	NA	23.6	7.70	11/07/2024	14:25
7	P4760-01	1	Water	NA	NA	24.6	8.27	11/07/2024	14:33
8	P4760-01DUP	1	Water	NA	NA	24.7	8.28	11/07/2024	14:34
9	P4764-02	1	Water	NA	NA	20.1	7.08	11/07/2024	14:35
10	CCV2	1	Water	NA	NA	20.3	12.02	11/07/2024	14:40

VB 133338

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p4764 WorkList ID : 185218 Department : Wet-Chemistry Date : 11-07-2024 13:45:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4759-01	C MONTCLAIR-TOTE-1	Water	pH	Cool 4 deg C	PSEG03	L23	11/07/2024	9040C
P4760-01	C LPA-TOTE-1	Water	pH	Cool 4 deg C	PSEG03	L23	11/07/2024	9040C
P4764-02	TW-WTS-01	Water	pH	Cool 4 deg C	ENTA05	L23	11/07/2024	9040C

Date/Time 11/07/24 13:50
Raw Sample Received by: JTCM
Raw Sample Relinquished by: JTCM

Date/Time 11/07/24 16:00
Raw Sample Received by: JTCM
Raw Sample Relinquished by: JTCM