



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4764

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133337

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	11/07/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	11/07/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	11/07/2024