

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: P4765

Project: Linden Generating Station - LRSA

RunNo.: LB133358

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/08/2024
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/08/2024
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/08/2024
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/08/2024
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/08/2024
Sample ID: CCB5 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	11/08/2024

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: P4765

Project: Linden Generating Station - LRSA

RunNo.: LB133431

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB COD	mg/L	< 5.0000	5.0000	U	2.35	10	10/14/2024
Sample ID: CCB1 COD	mg/L	< 5.0000	5.0000	U	2.35	10	11/13/2024
Sample ID: CCB2 COD	mg/L	< 5.0000	5.0000	U	2.35	10	11/13/2024
Sample ID: CCB3 COD	mg/L	< 5.0000	5.0000	U	2.35	10	11/13/2024

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: P4765

Project: Linden Generating Station - LRSA

RunNo.: LB133431

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------