

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: P4765

Project: Linden Generating Station - LRSA

Run No.: LB133342

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133342BS							
BOD5	mg/L	198	171		86	1	84.6-115.4	11/07/2024

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: P4765

Project: Linden Generating Station - LRSA

Run No.: LB133353

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133353BS							
TSS	mg/L	550	542		98	1	90-110	11/08/2024

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: P4765

Project: Linden Generating Station - LRSA

Run No.: LB133431

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133431BS							
COD	mg/L	50	50.3		101	1	90-110	11/13/2024

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: P4765

Project: Linden Generating Station - LRSA

Run No.: LB133358

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB164764BS							
Ammonia as N	mg/L	1	0.98		98	1	90-110	11/08/2024