

Surrogate Summary

SW-846

SDG No.: P4765

Client: Parkway Generation Operating LLC

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4765-02	DSN002	2-Methylnaphthalene-d10	0.4	0.35	86		20	139
		Fluoranthene-d10	0.4	0.42	105		30	150
		Nitrobenzene-d5	0.4	0.35	88		27	123
		2-Fluorobiphenyl	0.4	0.32	79		34	132
		Terphenyl-d14	0.4	0.43	107		35	157
P4765-06	DSN003	2-Methylnaphthalene-d10	0.4	0.30	76		20	139
		Fluoranthene-d10	0.4	0.37	93		30	150
		Nitrobenzene-d5	0.4	0.29	74		27	123
		2-Fluorobiphenyl	0.4	0.28	70		34	132
		Terphenyl-d14	0.4	0.40	99		35	157
PB164785BL	PB164785BL	2-Methylnaphthalene-d10	0.4	0.35	87		20	139
		Fluoranthene-d10	0.4	0.39	97		30	150
		Nitrobenzene-d5	0.4	0.38	95		27	123
		2-Fluorobiphenyl	0.4	0.37	91		34	132
		Terphenyl-d14	0.4	0.42	104		35	157
PB164785BS	PB164785BS	2-Methylnaphthalene-d10	0.4	0.44	110		20	139
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.36	90		27	123
		2-Fluorobiphenyl	0.4	0.37	93		34	132
		Terphenyl-d14	0.4	0.41	101		35	157
PB164785BSD	PB164785BSD	2-Methylnaphthalene-d10	0.4	0.37	93		20	139
		Fluoranthene-d10	0.4	0.29	72		30	150
		Nitrobenzene-d5	0.4	0.30	75		27	123
		2-Fluorobiphenyl	0.4	0.30	75		34	132
		Terphenyl-d14	0.4	0.32	79		35	157