

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PSEG04

Lab Code: CHEM

SAS No.: P4765

SDG NO.: P4765

Lab File ID: BN035061.D

DFTPP Injection Date: 11/13/2024

Instrument ID: BNA_N

DFTPP Injection Time: 12:01

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	19.6
68	Less than 2.0% of mass 69	0.4 (1.4) 1
69	Mass 69 relative abundance	29.1
70	Less than 2.0% of mass 69	0.2 (0.7) 1
127	10.0 - 80.0% of mass 198	35.9
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	27.6
365	Greater than 1% of mass 198	4.4
441	Present, but less than mass 443	9.3
442	Greater than 50% of mass 198	59.1
443	15.0 - 24.0% of mass 442	11.2 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN035062.D	11/13/2024	12:40
SSTDICC0.2	SSTDICC0.2	BN035063.D	11/13/2024	13:16
SSTDICCC0.4	SSTDICCC0.4	BN035064.D	11/13/2024	13:52
SSTDICC0.8	SSTDICC0.8	BN035065.D	11/13/2024	14:28
SSTDICC1.6	SSTDICC1.6	BN035066.D	11/13/2024	15:04
SSTDICC3.2	SSTDICC3.2	BN035067.D	11/13/2024	15:39
SSTDICC5.0	SSTDICC5.0	BN035068.D	11/13/2024	16:15
PB164785BL	PB164785BL	BN035070.D	11/13/2024	17:27
PB164785BS	PB164785BS	BN035071.D	11/13/2024	18:03
PB164785BSD	PB164785BSD	BN035072.D	11/13/2024	18:38
DSN002	P4765-02	BN035073.D	11/13/2024	19:14
DSN003	P4765-06	BN035074.D	11/13/2024	19:50