

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JPCL Engineering **SDG No.:** P4768
Contract: JPCL01 **Lab Code:** CHEM **Case No.:** P4768 **SAS No.:** P4768
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV82	Mercury	3.82	4.0	96	90 - 110	CV	11/12/2024	11:00	LB133402

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JPCL Engineering **SDG No.:** P4768
Contract: JPCL01 **Lab Code:** CHEM **Case No.:** P4768 **SAS No.:** P4768
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV71	Mercury	4.90	5.0	98	90 - 110	CV	11/12/2024	11:09	LB133402
CCV72	Mercury	5.04	5.0	101	90 - 110	CV	11/12/2024	11:39	LB133402
CCV73	Mercury	5.09	5.0	102	90 - 110	CV	11/12/2024	12:09	LB133402
CCV74	Mercury	5.22	5.0	104	90 - 110	CV	11/12/2024	12:36	LB133402

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JPCL Engineering **SDG No.:** P4768
Contract: JPCL01 **Lab Code:** CHEM **Case No.:** P4768 **SAS No.:** P4768
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1040	1000	104	90 - 110	P	11/14/2024	13:58	LB133457
	Barium	520	520	100	90 - 110	P	11/14/2024	13:58	LB133457
	Cadmium	508	510	100	90 - 110	P	11/14/2024	13:58	LB133457
	Chromium	533	520	102	90 - 110	P	11/14/2024	13:58	LB133457
	Lead	1020	1000	102	90 - 110	P	11/14/2024	13:58	LB133457
	Selenium	1040	1000	104	90 - 110	P	11/14/2024	13:58	LB133457
	Silver	255	250	102	90 - 110	P	11/14/2024	13:58	LB133457

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JPCL Engineering **SDG No.:** P4768
Contract: JPCL01 **Lab Code:** CHEM **Case No.:** P4768 **SAS No.:** P4768
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5030	5000	100	90 - 110	P	11/14/2024	15:22	LB133457
	Barium	10200	10000	102	90 - 110	P	11/14/2024	15:22	LB133457
	Cadmium	2490	2500	100	90 - 110	P	11/14/2024	15:22	LB133457
	Chromium	1020	1000	102	90 - 110	P	11/14/2024	15:22	LB133457
	Lead	4980	5000	100	90 - 110	P	11/14/2024	15:22	LB133457
	Selenium	5080	5000	102	90 - 110	P	11/14/2024	15:22	LB133457
	Silver	1260	1250	101	90 - 110	P	11/14/2024	15:22	LB133457
CCV02	Arsenic	4660	5000	93	90 - 110	P	11/14/2024	15:35	LB133457
	Barium	9380	10000	94	90 - 110	P	11/14/2024	15:35	LB133457
	Cadmium	2340	2500	94	90 - 110	P	11/14/2024	15:35	LB133457
	Chromium	952	1000	95	90 - 110	P	11/14/2024	15:35	LB133457
	Lead	4680	5000	94	90 - 110	P	11/14/2024	15:35	LB133457
	Selenium	4660	5000	93	90 - 110	P	11/14/2024	15:35	LB133457
	Silver	1180	1250	94	90 - 110	P	11/14/2024	15:35	LB133457
CCV03	Arsenic	5110	5000	102	90 - 110	P	11/14/2024	16:27	LB133457
	Barium	9600	10000	96	90 - 110	P	11/14/2024	16:27	LB133457
	Cadmium	2540	2500	102	90 - 110	P	11/14/2024	16:27	LB133457
	Chromium	1020	1000	102	90 - 110	P	11/14/2024	16:27	LB133457
	Lead	5090	5000	102	90 - 110	P	11/14/2024	16:27	LB133457
	Selenium	5130	5000	103	90 - 110	P	11/14/2024	16:27	LB133457
	Silver	1250	1250	100	90 - 110	P	11/14/2024	16:27	LB133457
CCV04	Arsenic	4730	5000	94	90 - 110	P	11/14/2024	17:23	LB133457
	Barium	9470	10000	95	90 - 110	P	11/14/2024	17:23	LB133457
	Cadmium	2380	2500	95	90 - 110	P	11/14/2024	17:23	LB133457
	Chromium	962	1000	96	90 - 110	P	11/14/2024	17:23	LB133457
	Lead	4770	5000	95	90 - 110	P	11/14/2024	17:23	LB133457
	Selenium	4710	5000	94	90 - 110	P	11/14/2024	17:23	LB133457
	Silver	1190	1250	95	90 - 110	P	11/14/2024	17:23	LB133457
CCV05	Arsenic	4900	5000	98	90 - 110	P	11/14/2024	18:12	LB133457
	Barium	9970	10000	100	90 - 110	P	11/14/2024	18:12	LB133457
	Cadmium	2470	2500	99	90 - 110	P	11/14/2024	18:12	LB133457
	Chromium	996	1000	100	90 - 110	P	11/14/2024	18:12	LB133457
	Lead	4960	5000	99	90 - 110	P	11/14/2024	18:12	LB133457
	Selenium	4860	5000	97	90 - 110	P	11/14/2024	18:12	LB133457

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JPCL Engineering **SDG No.:** P4768

Contract: JPCL01 **Lab Code:** CHEM **Case No.:** P4768 **SAS No.:** P4768

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1240	1250	99	90 - 110	P	11/14/2024	18:12	LB133457
CCV06	Arsenic	4770	5000	95	90 - 110	P	11/14/2024	19:03	LB133457
	Barium	9920	10000	99	90 - 110	P	11/14/2024	19:03	LB133457
	Cadmium	2400	2500	96	90 - 110	P	11/14/2024	19:03	LB133457
	Chromium	971	1000	97	90 - 110	P	11/14/2024	19:03	LB133457
	Lead	4820	5000	96	90 - 110	P	11/14/2024	19:03	LB133457
	Selenium	4720	5000	94	90 - 110	P	11/14/2024	19:03	LB133457
	Silver	1210	1250	97	90 - 110	P	11/14/2024	19:03	LB133457
CCV07	Arsenic	4840	5000	97	90 - 110	P	11/14/2024	19:55	LB133457
	Barium	9800	10000	98	90 - 110	P	11/14/2024	19:55	LB133457
	Cadmium	2420	2500	97	90 - 110	P	11/14/2024	19:55	LB133457
	Chromium	980	1000	98	90 - 110	P	11/14/2024	19:55	LB133457
	Lead	4850	5000	97	90 - 110	P	11/14/2024	19:55	LB133457
	Selenium	4850	5000	97	90 - 110	P	11/14/2024	19:55	LB133457
	Silver	1220	1250	97	90 - 110	P	11/14/2024	19:55	LB133457
CCV08	Arsenic	4920	5000	98	90 - 110	P	11/14/2024	20:47	LB133457
	Barium	9680	10000	97	90 - 110	P	11/14/2024	20:47	LB133457
	Cadmium	2450	2500	98	90 - 110	P	11/14/2024	20:47	LB133457
	Chromium	1000	1000	100	90 - 110	P	11/14/2024	20:47	LB133457
	Lead	4900	5000	98	90 - 110	P	11/14/2024	20:47	LB133457
	Selenium	4960	5000	99	90 - 110	P	11/14/2024	20:47	LB133457
	Silver	1230	1250	99	90 - 110	P	11/14/2024	20:47	LB133457
CCV09	Arsenic	4790	5000	96	90 - 110	P	11/14/2024	22:04	LB133457
	Barium	9850	10000	98	90 - 110	P	11/14/2024	22:04	LB133457
	Cadmium	2420	2500	97	90 - 110	P	11/14/2024	22:04	LB133457
	Chromium	980	1000	98	90 - 110	P	11/14/2024	22:04	LB133457
	Lead	4830	5000	97	90 - 110	P	11/14/2024	22:04	LB133457
	Selenium	4760	5000	95	90 - 110	P	11/14/2024	22:04	LB133457
	Silver	1210	1250	97	90 - 110	P	11/14/2024	22:04	LB133457
CCV10	Arsenic	4790	5000	96	90 - 110	P	11/14/2024	22:58	LB133457
	Barium	9850	10000	98	90 - 110	P	11/14/2024	22:58	LB133457
	Cadmium	2400	2500	96	90 - 110	P	11/14/2024	22:58	LB133457
	Chromium	974	1000	97	90 - 110	P	11/14/2024	22:58	LB133457
	Lead	4800	5000	96	90 - 110	P	11/14/2024	22:58	LB133457

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JPCL Engineering **SDG No.:** P4768
Contract: JPCL01 **Lab Code:** CHEM **Case No.:** P4768 **SAS No.:** P4768
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Selenium	4740	5000	95	90 - 110	P	11/14/2024	22:58	LB133457
	Silver	1220	1250	97	90 - 110	P	11/14/2024	22:58	LB133457