



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>JPCL Engineering</u>				SDG No.:	<u>P4768</u>			
Contract:	<u>JPCL01</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>P4768</u>	SAS No.:	<u>P4768</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB82	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	11:06	LB133402

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JPCL Engineering **SDG No.:** P4768
Contract: JPCL01 **Lab Code:** CHEM **Case No.:** P4768 **SAS No.:** P4768

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB71	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	11:11	LB133402
CCB72	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	11:41	LB133402
CCB73	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	12:11	LB133402
CCB74	Mercury	0.20	+/-0.20	U	0.20	CV	11/12/2024	12:39	LB133402

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	JPCL Engineering				SDG No.: P4768				
Contract:	JPCL01	Lab Code:	CHEM	Case No.: P4768		SAS No.: P4768			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	14:36	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	14:36	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	14:36	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	14:36	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	14:36	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	14:36	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	14:36	LB133457

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JPCL Engineering		SDG No.: P4768							
Contract: JPCL01	Lab Code: CHEM	Case No.: P4768	SAS No.: P4768						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	15:27	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	15:27	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	15:27	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	15:27	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	15:27	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	15:27	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	15:27	LB133457
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	15:40	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	15:40	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	15:40	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	15:40	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	15:40	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	15:40	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	15:40	LB133457
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	16:31	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	16:31	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	16:31	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	16:31	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	16:31	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	16:31	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	16:31	LB133457
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	17:27	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	17:27	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	17:27	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	17:27	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	17:27	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	17:27	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	17:27	LB133457
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	18:17	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	18:17	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	18:17	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	18:17	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	18:17	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	18:17	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	18:17	LB133457
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	19:07	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	19:07	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	19:07	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	19:07	LB133457

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	JPCL Engineering			SDG No.:	P4768				
Contract:	JPCL01	Lab Code:	CHEM	Case No.:	P4768	SAS No.:	P4768		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	19:07	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	19:07	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	19:07	LB133457
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	19:59	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	19:59	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	19:59	LB133457
CCB08	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	19:59	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	19:59	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	19:59	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	19:59	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	21:42	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	21:42	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	21:42	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	21:42	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	21:42	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	21:42	LB133457
CCB09	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	21:42	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	22:08	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	22:08	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	22:08	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	22:08	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	22:08	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	22:08	LB133457
CCB10	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	22:08	LB133457
	Arsenic	20.0	+/-20.0	U	20.0	P	11/14/2024	23:03	LB133457
	Barium	100	+/-100	U	100	P	11/14/2024	23:03	LB133457
	Cadmium	6.00	+/-6.00	U	6.00	P	11/14/2024	23:03	LB133457
	Chromium	10.0	+/-10.0	U	10.0	P	11/14/2024	23:03	LB133457
	Lead	12.0	+/-12.0	U	12.0	P	11/14/2024	23:03	LB133457
	Selenium	20.0	+/-20.0	U	20.0	P	11/14/2024	23:03	LB133457
	Silver	10.0	+/-10.0	U	10.0	P	11/14/2024	23:03	LB133457