

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering **SDG No.:** P4768
Contract: JPCL01 **Lab Code:** CHEM **Case No.:** P4768 **SAS No.:** P4768

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164797BS							
Arsenic	mg/Kg	36.5	35.8		98	80 - 120	P
Barium	mg/Kg	9.1	8.87		98	80 - 120	P
Cadmium	mg/Kg	9.1	8.62		95	80 - 120	P
Chromium	mg/Kg	18.3	18.1		99	80 - 120	P
Lead	mg/Kg	45.7	43.9		96	80 - 120	P
Selenium	mg/Kg	91.3	88.2		97	80 - 120	P
Silver	mg/Kg	3.4	3.31		97	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering

SDG No.: P4768

Contract: JPCL01

Lab Code: CHEM

Case No.: P4768

SAS No.: P4768

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164891BS Mercury	mg/Kg	0.26	0.30		114	80 - 120	CV