

Continuing Calibration Report for SequenceID : FC111324AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067734.D

Aliphatic C9-C12	7176964.000	60.000	3.169	6.447	119616.067	135469.641	11.703
Aliphatic C12-C16	4748255.000	40.000	6.448	9.836	118706.375	131681.730	9.854
Aliphatic C16-C21	6999820.000	60.000	9.837	13.192	116663.667	127235.092	8.309
Aliphatic C21-C28	9375270.000	80.000	13.193	16.845	117190.875	120090.398	2.414
Aliphatic C28-C40	13486380.000	120.000	16.846	21.681	112386.500	101093.195	-11.171
Aliphatic EPH	41786689.000	360.000	3.169	21.681	116074.136	118799.912	2.294