

Continuing Calibration Report for SequenceID : FF111424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015026.D

Aromatic C10-C12	5732257.000	40.000	4.508	6.173	143306.425	154569.237	7.287
Aromatic C12-C16	8765756.000	60.000	6.174	9.065	146095.933	153560.844	4.861
Aromatic C16-C21	10862559.000	80.000	9.066	13.377	135781.988	142096.237	4.444
Aromatic C21-C36	22777506.000	180.000	13.378	18.821	126541.700	138041.042	8.330
Aromatic EPH	48138078.000	360.000	4.508	18.821	133716.883	143365.296	6.730