

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** JPCL Engineering
Lab Code: CHEM **Cas No:** P4768 **SAS No :** P4768 **SDG No:** P4768
Sample No : P4768-07MS **Datafile:** FC067740.D
Client ID : B-6-3MS

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10.5	0.42	7.83	71		(40-140)
Aliphatic C12-C16	7	0.53	7.46	99		(40-140)
Aliphatic C16-C21	10.5	0.40	11.7	107		(40-140)
Aliphatic C21-C28	14.0	0	15.8	113		(40-140)
Aliphatic C28-C40	21.0	6.13	36.3	144	*	(40-140)