

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** JPCL Engineering
Lab Code: CHEM **Cas No:** P4768 **SAS No :** P4768 **SDG No:** P4768
Sample No : P4768-07MS **Datafile:** FD048721.D
Client ID : B-6-3MS

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aromatic C10-C12	7	0	5.58	80		(40-140)
Aromatic C12-C16	10.5	0.76	10.3	91		(40-140)
Aromatic C16-C21	17.5	11.0	28.3	99		(40-140)
Aromatic C21-C36	27.9	13.1	43.5	108		(40-140)