

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** JPCL Engineering
Lab Code: CHEM **Cas No:** P4768 **SAS No :** P4768 **SDG No:** P4768
Sample No : P4768-07MSD **Datafile:** FD048722.D
Client ID : B-6-3MSD

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS QC Limit Of RPD
Aromatic C10-C12	7	0	5.35	77		4.09 (40-140) 25
Aromatic C12-C16	10.5	0.76	12.9	116		24 (40-140) 25
Aromatic C16-C21	17.4	11.0	27.1	92		7 (40-140) 25
Aromatic C21-C36	27.9	13.1	41.7	102		5.71 (40-140) 25