

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** JPCL Engineering
Lab Code: CHEM **Cas No:** P4768 **SAS No :** P4768 **SDG No:** P4768
Sample No : PB164943BS **Datafile:** FD048717.D
Client ID : PB164943BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aromatic C10-C12	6.7	4.64	70		(40-140)
Aromatic C12-C16	10.0	8.32	83		(40-140)
Aromatic C16-C21	16.7	16.9	101		(40-140)
Aromatic C21-C36	26.7	26.2	98		(40-140)