

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/08/2024

Lab Sample No.: AR1660CCC500 Data File : PO107803.D Time Analyzed: 09:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.528	5.423	5.623	512.590	500.000	2.5
Aroclor-1016-2	5.551	5.445	5.645	497.300	500.000	-0.5
Aroclor-1016-3	5.613	5.507	5.707	488.250	500.000	-2.4
Aroclor-1016-4	5.711	5.604	5.804	509.660	500.000	1.9
Aroclor-1016-5	6.006	5.899	6.099	510.800	500.000	2.2
Aroclor-1260-1	7.136	7.026	7.226	478.840	500.000	-4.2
Aroclor-1260-2	7.394	7.282	7.482	509.330	500.000	1.9
Aroclor-1260-3	7.757	7.643	7.843	522.220	500.000	4.4
Aroclor-1260-4	7.981	7.868	8.068	546.580	500.000	9.3
Aroclor-1260-5	8.296	8.181	8.381	579.970	500.000	16.0
Decachlorobiphenyl	10.086	9.958	10.158	47.010	50.000	-6.0
Tetrachloro-m-xylene	4.375	4.274	4.474	52.120	50.000	4.2

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/08/2024

Lab Sample No.: AR1660CCC500 Data File : PO107803.D Time Analyzed: 09:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	533.610	500.000	6.7
Aroclor-1016-2	4.742	4.645	4.845	557.880	500.000	11.6
Aroclor-1016-3	4.918	4.821	5.021	535.090	500.000	7.0
Aroclor-1016-4	4.959	4.862	5.062	495.800	500.000	-0.8
Aroclor-1016-5	5.173	5.076	5.276	529.230	500.000	5.8
Aroclor-1260-1	6.205	6.107	6.307	509.290	500.000	1.9
Aroclor-1260-2	6.392	6.294	6.494	536.270	500.000	7.3
Aroclor-1260-3	6.545	6.448	6.648	524.930	500.000	5.0
Aroclor-1260-4	7.017	6.918	7.118	519.400	500.000	3.9
Aroclor-1260-5	7.258	7.159	7.359	547.840	500.000	9.6
Decachlorobiphenyl	8.641	8.539	8.739	51.420	50.000	2.8
Tetrachloro-m-xylene	3.642	3.545	3.745	54.950	50.000	9.9