

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/08/2024

Lab Sample No.: AR1660CCC500 Data File : PO107816.D Time Analyzed: 16:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.423	5.623	499.460	500.000	-0.1
Aroclor-1016-2	5.548	5.445	5.645	494.190	500.000	-1.2
Aroclor-1016-3	5.610	5.507	5.707	494.020	500.000	-1.2
Aroclor-1016-4	5.708	5.604	5.804	518.640	500.000	3.7
Aroclor-1016-5	6.003	5.899	6.099	506.230	500.000	1.2
Aroclor-1260-1	7.133	7.026	7.226	473.640	500.000	-5.3
Aroclor-1260-2	7.391	7.282	7.482	493.530	500.000	-1.3
Aroclor-1260-3	7.753	7.643	7.843	498.890	500.000	-0.2
Aroclor-1260-4	7.979	7.868	8.068	521.870	500.000	4.4
Aroclor-1260-5	8.294	8.181	8.381	533.910	500.000	6.8
Decachlorobiphenyl	10.083	9.958	10.158	47.830	50.000	-4.3
Tetrachloro-m-xylene	4.372	4.274	4.474	53.040	50.000	6.1

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/08/2024

Lab Sample No.: AR1660CCC500 Data File : PO107816.D Time Analyzed: 16:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	546.750	500.000	9.4
Aroclor-1016-2	4.742	4.645	4.845	561.600	500.000	12.3
Aroclor-1016-3	4.918	4.821	5.021	538.410	500.000	7.7
Aroclor-1016-4	4.960	4.862	5.062	491.310	500.000	-1.7
Aroclor-1016-5	5.173	5.076	5.276	529.370	500.000	5.9
Aroclor-1260-1	6.205	6.107	6.307	498.090	500.000	-0.4
Aroclor-1260-2	6.392	6.294	6.494	522.550	500.000	4.5
Aroclor-1260-3	6.546	6.448	6.648	501.750	500.000	0.4
Aroclor-1260-4	7.016	6.918	7.118	481.780	500.000	-3.6
Aroclor-1260-5	7.258	7.159	7.359	503.170	500.000	0.6
Decachlorobiphenyl	8.640	8.539	8.739	46.970	50.000	-6.1
Tetrachloro-m-xylene	3.642	3.545	3.745	55.210	50.000	10.4