

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/08/2024

Lab Sample No.: AR1660CCC500 Data File : PO107826.D Time Analyzed: 20:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.423	5.623	537.430	500.000	7.5
Aroclor-1016-2	5.547	5.445	5.645	524.340	500.000	4.9
Aroclor-1016-3	5.610	5.507	5.707	518.740	500.000	3.7
Aroclor-1016-4	5.707	5.604	5.804	541.990	500.000	8.4
Aroclor-1016-5	6.003	5.899	6.099	543.600	500.000	8.7
Aroclor-1260-1	7.133	7.026	7.226	512.570	500.000	2.5
Aroclor-1260-2	7.391	7.282	7.482	542.700	500.000	8.5
Aroclor-1260-3	7.754	7.643	7.843	564.410	500.000	12.9
Aroclor-1260-4	7.978	7.868	8.068	583.640	500.000	16.7
Aroclor-1260-5	8.293	8.181	8.381	563.150	500.000	12.6
Decachlorobiphenyl	10.082	9.958	10.158	48.870	50.000	-2.3
Tetrachloro-m-xylene	4.373	4.274	4.474	54.750	50.000	9.5

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/08/2024

Lab Sample No.: AR1660CCC500 Data File : PO107826.D Time Analyzed: 20:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	547.220	500.000	9.4
Aroclor-1016-2	4.742	4.645	4.845	566.890	500.000	13.4
Aroclor-1016-3	4.918	4.821	5.021	546.630	500.000	9.3
Aroclor-1016-4	4.959	4.862	5.062	500.500	500.000	0.1
Aroclor-1016-5	5.172	5.076	5.276	537.790	500.000	7.6
Aroclor-1260-1	6.204	6.107	6.307	523.810	500.000	4.8
Aroclor-1260-2	6.391	6.294	6.494	552.330	500.000	10.5
Aroclor-1260-3	6.545	6.448	6.648	537.660	500.000	7.5
Aroclor-1260-4	7.015	6.918	7.118	530.390	500.000	6.1
Aroclor-1260-5	7.256	7.159	7.359	558.910	500.000	11.8
Decachlorobiphenyl	8.638	8.539	8.739	50.630	50.000	1.3
Tetrachloro-m-xylene	3.642	3.545	3.745	55.520	50.000	11.0