

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JPCL01

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG No.: P4768

Instrument ID: BNA_F Calibration Date/Time: 11/14/2024 17:02

Lab File ID: BF140366.D Init. Calib. Date(s): 11/13/2024 11/13/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:01 12:48

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.171	1.188		1.5	
Benzaldehyde	0.951	0.861		-9.5	
Phenol-d6	1.587	1.578		-0.6	
Phenol	1.674	1.678		0.2	20.0
bis(2-Chloroethyl)ether	1.268	1.243		-2.0	
2-Chlorophenol	1.258	1.256		-0.2	
2-Methylphenol	1.035	1.033		-0.2	
2,2-oxybis(1-Chloropropane)	1.752	1.678		-4.2	
Acetophenone	0.465	0.463		-0.4	
3+4-Methylphenols	1.295	1.288		-0.5	
n-Nitroso-di-n-propylamine	0.959	0.909	0.050	-5.2	
Nitrobenzene-d5	0.384	0.381		-0.8	
Hexachloroethane	0.520	0.504		-3.1	
Nitrobenzene	0.400	0.401		0.3	
Isophorone	0.680	0.661		-2.8	
2-Nitrophenol	0.177	0.178		0.6	20.0
2,4-Dimethylphenol	0.227	0.227		0.0	
bis(2-Chloroethoxy)methane	0.417	0.404		-3.1	
2,4-Dichlorophenol	0.281	0.281		0.0	20.0
Naphthalene	1.015	1.009		-0.6	
4-Chloroaniline	0.353	0.327		-7.4	
Hexachlorobutadiene	0.199	0.195		-2.0	20.0
Caprolactam	0.093	0.094		1.1	
4-Chloro-3-methylphenol	0.311	0.311		0.0	20.0
2-Methylnaphthalene	0.651	0.642		-1.4	
Hexachlorocyclopentadiene	0.142	0.133	0.050	-6.3	
2,4,6-Trichlorophenol	0.363	0.368		1.4	20.0
2-Fluorobiphenyl	1.244	1.231		-1.0	
2,4,5-Trichlorophenol	0.397	0.409		3.0	
1,1-Biphenyl	1.455	1.467		0.8	
2-Chloronaphthalene	1.108	1.111		0.3	
2-Nitroaniline	0.368	0.374		1.6	
Dimethylphthalate	1.304	1.272		-2.5	
Acenaphthylene	1.677	1.672		-0.3	
2,6-Dinitrotoluene	0.297	0.303		2.0	
3-Nitroaniline	0.312	0.313		0.3	
Acenaphthene	1.133	1.135		0.2	20.0
2,4-Dinitrophenol	0.153	0.154	0.050	0.7	
4-Nitrophenol	0.228	0.236	0.050	3.5	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JPCL01

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG No.: P4768

Instrument ID: BNA_F Calibration Date/Time: 11/14/2024 17:02

Lab File ID: BF140366.D Init. Calib. Date(s): 11/13/2024 11/13/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:01 12:48

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.603	1.616		0.8	
2,4-Dinitrotoluene	0.391	0.397		1.5	
Diethylphthalate	1.333	1.281		-3.9	
4-Chlorophenyl-phenylether	0.625	0.604		-3.4	
Fluorene	1.267	1.256		-0.9	
4-Nitroaniline	0.319	0.328		2.8	
4,6-Dinitro-2-methylphenol	0.108	0.114		5.6	
n-Nitrosodiphenylamine	0.578	0.569		-1.6	20.0
2,4,6-Tribromophenol	0.199	0.200		0.5	
4-Bromophenyl-phenylether	0.200	0.193		-3.5	
Hexachlorobenzene	0.225	0.223		-0.9	
Atrazine	0.175	0.164		-6.3	
Pentachlorophenol	0.121	0.124		2.5	20.0
Phenanthrene	0.940	0.926		-1.5	
Anthracene	0.921	0.905		-1.7	
Carbazole	0.909	0.912		0.3	
Di-n-butylphthalate	1.091	1.021		-6.4	
Fluoranthene	1.054	1.029		-2.4	20.0
Pyrene	1.711	1.783		4.2	
Terphenyl-d14	1.152	1.148		-0.3	
Butylbenzylphthalate	0.657	0.642		-2.3	
3,3-Dichlorobenzidine	0.418	0.409		-2.2	
Benzo (a) anthracene	1.314	1.308		-0.5	
Chrysene	1.199	1.156		-3.6	
Bis(2-ethylhexyl)phthalate	0.877	0.791		-9.8	
Di-n-octyl phthalate	1.235	1.083		-12.3	20.0
Benzo (b) fluoranthene	1.308	1.240		-5.2	
Benzo (k) fluoranthene	1.069	1.043		-2.4	
Benzo (a) pyrene	1.016	1.000		-1.6	20.0
Indeno (1,2,3-cd) pyrene	1.235	1.295		4.9	
Dibenzo (a,h) anthracene	1.021	1.058		3.6	
Benzo (g,h,i) perylene	1.044	1.108		6.1	
1,2,4,5-Tetrachlorobenzene	0.553	0.555		0.4	
1,4-Dioxane	0.522	0.547		4.8	20.0
2,3,4,6-Tetrachlorophenol	0.313	0.320		2.2	

All other compounds must meet a minimum RRF of 0.010.