

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110524

Review By	yogesh	Review On	11/6/2024 3:09:46 AM
Supervise By	mohammad	Supervise On	11/6/2024 3:16:35 AM
SubDirectory	BF110524	HP Acquire Method	BNA_F
		HP Processing Method	bf110524

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12324,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140230.D	05 Nov 2024 11:56		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140231.D	05 Nov 2024 12:26		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140232.D	05 Nov 2024 12:52	Compound#32,54,65 Removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF140233.D	05 Nov 2024 13:18	Compound#54 Kept on LR	RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF140234.D	05 Nov 2024 13:45		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF140235.D	05 Nov 2024 14:11	Calibration Good for DOD & 625.1 except for Benzidine	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF140236.D	05 Nov 2024 14:37	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF140237.D	05 Nov 2024 15:03		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF140238.D	05 Nov 2024 15:30		RC/JU	Ok,M
10	SSTDICV040	ICVBF110524	BF140239.D	05 Nov 2024 16:07		RC/JU	Ok
11	PB164366BL	PB164366BL	BF140240.D	05 Nov 2024 16:37		RC/JU	Ok
12	P4368-02	LOQ-SOIL-02-QT4-202	BF140241.D	05 Nov 2024 17:03	LOQ-SOIL-5 ppm	RC/JU	Ok
13	P4368-02	LOQ-SOIL-02-QT4-202	BF140242.D	05 Nov 2024 17:30	LOQ-SOIL-10 PPM (Used for All compounds except compounds#54,77)	RC/JU	Ok,M
14	P4368-08	LOQ-WATER-02-QT4-2	BF140243.D	05 Nov 2024 17:56	LOQ-Water-5 ppm	RC/JU	Ok,M
15	P4368-02	LOQ-SOIL-02-QT4-202	BF140244.D	05 Nov 2024 18:22	LOQ-SOIL-5 ppm, Already Analyzed	RC/JU	Not Ok
16	P4368-08	LOQ-WATER-02-QT4-2	BF140245.D	05 Nov 2024 18:49	LOQ-Water-10 ppm, Internal Standard Fail	RC/JU	Not Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110524

Review By	yogesh	Review On	11/6/2024 3:09:46 AM		
Supervise By	mohammad	Supervise On	11/6/2024 3:16:35 AM		
SubDirectory	BF110524	HP Acquire Method	BNA_F	HP Processing Method	bf110524
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12324,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4368-01	LOD-MDL-SOIL-01-QT	BF140246.D	05 Nov 2024 19:15	LOD-MDL-SOIL- 4 ppm	RC/JU	Ok
18	P4368-01	LOD-MDL-SOIL-01-QT	BF140247.D	05 Nov 2024 19:41	LOD-MDL-SOIL- 8 ppm	RC/JU	Ok,M
19	P4368-07	LOD-MDL-WATER-01-QT	BF140248.D	05 Nov 2024 20:07	LOD-MDL-WATER- 4 ppm	RC/JU	Ok,M
20	P4368-07	LOD-MDL-WATER-01-QT	BF140249.D	05 Nov 2024 20:34	LOD-MDL-WATER- 8 ppm	RC/JU	Ok,M
21	PB164369BL	PB164369BL	BF140250.D	05 Nov 2024 21:00		RC/JU	Ok
22	SSTDCCC040	SSTDCCC040EC	BF140251.D	05 Nov 2024 21:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110824

Review By	yogesh	Review On	11/10/2024 11:36:46 PM
Supervise By	mohammad	Supervise On	11/11/2024 12:24:00 AM
SubDirectory	BF110824	HP Acquire Method	BNA_F
		HP Processing Method	bf110524

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12326,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140285.D	08 Nov 2024 09:09		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140286.D	08 Nov 2024 09:35		RC/JU	Ok,M
3	PB164765BL	PB164765BL	BF140287.D	08 Nov 2024 10:02		RC/JU	Ok
4	PB164750BS	PB164750BS	BF140288.D	08 Nov 2024 10:28		RC/JU	Ok,M
5	PB164680BL	PB164680BL	BF140289.D	08 Nov 2024 10:55		RC/JU	Ok
6	PB164680BS	PB164680BS	BF140290.D	08 Nov 2024 11:22		RC/JU	Ok,M
7	P4732-01	PPE-COMP	BF140291.D	08 Nov 2024 11:54	Internal standard fail	RC/JU	Ok
8	P4694-05RE	Z-04RE	BF140292.D	08 Nov 2024 12:21	Internal standard fail	RC/JU	Confirms
9	P4720-01DL	JC-701-COMP-01DL	BF140293.D	08 Nov 2024 12:47		RC/JU	Ok
10	P4695-01DL	Z-01DL	BF140294.D	08 Nov 2024 13:13		RC/JU	Ok,M
11	P4739-01	TP-14	BF140295.D	08 Nov 2024 13:40	Internal standard fail	RC/JU	Ok
12	P4694-01DL	Z-03ADL	BF140296.D	08 Nov 2024 14:06	Internal standard fail	RC/JU	Ok,M
13	P4739-05	BP-G2	BF140297.D	08 Nov 2024 14:32	Internal standard fail	RC/JU	Ok
14	P4739-09	BP-B2	BF140298.D	08 Nov 2024 14:58	Internal Standard Fail, Need 5X Dilution	RC/JU	Dilution
15	P4737-01	TP2	BF140299.D	08 Nov 2024 15:24	Internal standard fail	RC/JU	Ok
16	P4737-01MS	TP2MS	BF140300.D	08 Nov 2024 15:51	Internal standard fail	RC/JU	Ok
17	P4737-01MSD	TP2MSD	BF140301.D	08 Nov 2024 16:17	Internal standard fail	RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110824

Review By	yogesh	Review On	11/10/2024 11:36:46 PM		
Supervise By	mohammad	Supervise On	11/11/2024 12:24:00 AM		
SubDirectory	BF110824	HP Acquire Method	BNA_F	HP Processing Method	bf110524
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12326,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P4738-01	72-12018	BF140302.D	08 Nov 2024 16:43	Internal standard fail	RC/JU	ReRun
19	P4722-08	WC-2(0-6)	BF140303.D	08 Nov 2024 17:10	Internal standard fail, Need 2X Dilution	RC/JU	Dilution
20	P4739-13	TP-11	BF140304.D	08 Nov 2024 17:36	Internal Standard Fail	RC/JU	Ok,M
21	P4722-03	WC-1(0-6)	BF140305.D	08 Nov 2024 18:03	Internal Standard Fail	RC/JU	Ok,M
22	P4756-01	BP-B4	BF140306.D	08 Nov 2024 18:29	Internal Standard Fail	RC/JU	Ok
23	P4756-01MS	BP-B4MS	BF140307.D	08 Nov 2024 18:55	Internal Standard Fail	RC/JU	Ok
24	P4756-01MSD	BP-B4MSD	BF140308.D	08 Nov 2024 19:22	Internal Standard Fail	RC/JU	Ok
25	P4768-01	B-3-1	BF140309.D	08 Nov 2024 19:48	Internal Standard Fail	RC/JU	Ok
26	P4768-06	B-6-2	BF140310.D	08 Nov 2024 20:14	Internal standard fail, Need 2X Dilution	RC/JU	Dilution

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111324

Review By	yogesh	Review On	11/14/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:17 AM
SubDirectory	BF111324	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140331.D	13 Nov 2024 08:35		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140332.D	13 Nov 2024 09:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140333.D	13 Nov 2024 09:27	Compound#41,54,77 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF140334.D	13 Nov 2024 09:53		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF140335.D	13 Nov 2024 10:29		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF140336.D	13 Nov 2024 10:55	Injection Error	RC/JU	Not Ok
7	SSTDICC050	SSTDICC050	BF140337.D	13 Nov 2024 11:21		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF140338.D	13 Nov 2024 11:47		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF140339.D	13 Nov 2024 12:13	Compound#09 removed from 80 ppm	RC/JU	Ok,M
10	SSTDICCC040	SSTDICCC040	BF140340.D	13 Nov 2024 12:48		RC/JU	Ok
11	SSTDICV040	ICVBF111324	BF140341.D	13 Nov 2024 13:19		RC/JU	Ok
12	PB164401BL	PB164401BL	BF140342.D	13 Nov 2024 13:45		RC/JU	Ok
13	P4495-11	PT-BNA-SOIL	BF140343.D	13 Nov 2024 15:25	PT Sample, Need 5X Dilution	RC/JU	Dilution
14	P4495-11DL	PT-BNA-SOILDL	BF140344.D	13 Nov 2024 16:02		RC/JU	Ok
15	P3845-03	PT-BN-WP	BF140345.D	13 Nov 2024 16:40	PT Sample Analyzed for confirmation - Need 5X Dilution.	RC/JU	Dilution
16	P3845-03DL	PT-BN-WPDL	BF140346.D	13 Nov 2024 17:06		RC/JU	Ok
17	P3845-06	PT-ACIDS-WP	BF140347.D	13 Nov 2024 17:32	PT Sample Analyzed for confirmation - Need 5X Dilution.	RC/JU	Dilution

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111324

Review By	yogesh	Review On	11/14/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:17 AM
SubDirectory	BF111324	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	P3845-06DL	PT-ACIDS-WPDL	BF140348.D	13 Nov 2024 17:58	Dilution not matching (Low results)	RC/JU	Not Ok
19	SP6681	SP6681	BF140349.D	13 Nov 2024 18:39	8270-625.1 SPIKE SOLUTION	RC/JU	Ok,NR

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111424

Review By	yogesh	Review On	11/15/2024 5:15:55 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:37 AM
SubDirectory	BF111424	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140350.D	14 Nov 2024 08:47		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140351.D	14 Nov 2024 09:42		RC/JU	Ok
3	PB164908BL	PB164908BL	BF140352.D	14 Nov 2024 10:08		RC/JU	Ok
4	PB164911BS	PB164911BS	BF140353.D	14 Nov 2024 10:34		RC/JU	Ok,M
5	PB164911BSD	PB164911BSD	BF140354.D	14 Nov 2024 11:01		RC/JU	Ok,M
6	PB164911BL	PB164911BL	BF140355.D	14 Nov 2024 11:27		RC/JU	Ok
7	P3845-06DL	PT-ACIDS-WPDL	BF140356.D	14 Nov 2024 11:54		RC/JU	Ok
8	P4762-02	EFF-WW	BF140357.D	14 Nov 2024 13:03		RC/JU	Ok
9	P4792-03	OILY-STONE-DRUM	BF140358.D	14 Nov 2024 13:29		RC/JU	Ok
10	P4837-03	3TRK-STONE	BF140359.D	14 Nov 2024 13:55		RC/JU	Ok
11	P4834-03	SLP-1-STONE	BF140360.D	14 Nov 2024 14:21		RC/JU	Ok
12	P4779-01	SOIL-01	BF140361.D	14 Nov 2024 14:47		RC/JU	Ok
13	P4838-03	T1-STONE	BF140362.D	14 Nov 2024 15:13		RC/JU	Ok
14	P4718-04	WB-307-SW	BF140363.D	14 Nov 2024 15:39		RC/JU	Ok
15	P4812-01	RBR200035	BF140364.D	14 Nov 2024 16:04		RC/JU	Ok
16	DFTPP	DFTPP	BF140365.D	14 Nov 2024 16:35		RC/JU	Ok
17	SSTDCCC040	SSTDCCC040	BF140366.D	14 Nov 2024 17:02		RC/JU	Ok
18	PB164846BL	PB164846BL	BF140367.D	14 Nov 2024 17:28		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111424

Review By	yogesh	Review On	11/15/2024 5:15:55 AM				
Supervise By	mohammad	Supervise On	11/15/2024 5:49:37 AM				
SubDirectory	BF111424	HP Acquire Method	BNA_F	HP Processing Method	Bf111324		
STD. NAME		STD REF.#					
Tune/Reschk		SP6573					
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621					
CCC		SP6624					
Internal Standard/PEM		S12327,10ul/1000ul sample					
ICV/I.BLK		SP6559					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

19	P4722-10	WC-2(0-6)	BF140368.D	14 Nov 2024 17:59		RC/JU	Ok
20	P4722-15	WC-3(0-6)	BF140369.D	14 Nov 2024 18:25		RC/JU	Ok
21	P4791-02	HINCHMAN-OILY-WAT	BF140370.D	14 Nov 2024 18:51	Internal Standard Fail	RC/JU	Ok
22	P4792-04	MANHOLE-WASTE-DR	BF140371.D	14 Nov 2024 19:17		RC/JU	Ok
23	P4788-01	BP-G3	BF140372.D	14 Nov 2024 19:43		RC/JU	Ok
24	P4788-01MS	BP-G3MS	BF140373.D	14 Nov 2024 20:09		RC/JU	Ok
25	P4788-01MSD	BP-G3MSD	BF140374.D	14 Nov 2024 20:35		RC/JU	Ok
26	P4796-05RE	TP-3RE	BF140375.D	14 Nov 2024 21:01	Fax is already given	RC/JU	Confirms
27	P4796-01RE	TP-1RE	BF140376.D	14 Nov 2024 21:28	Fax is already given	RC/JU	Confirms
28	P4722-08DL	WC-2(0-6)DL	BF140377.D	14 Nov 2024 21:54		RC/JU	Ok,M
29	P4768-06DL	B-6-2DL	BF140378.D	14 Nov 2024 22:20		RC/JU	Ok,M
30	P4739-09DL	BP-B2DL	BF140379.D	14 Nov 2024 22:47		RC/JU	Ok
31	P4795-01RE	LAW-23-00189RE	BF140380.D	14 Nov 2024 23:13	Fax is already given	RC/JU	Confirms
32	P4811-01RE	RB24008RE	BF140381.D	14 Nov 2024 23:39	Internal Standard Fail	RC/JU	Confirms
33	P4816-01RE	331RE	BF140382.D	15 Nov 2024 00:05	Fax is already given	RC/JU	Confirms
34	P4807-05	BF-F14	BF140383.D	15 Nov 2024 00:32		RC/JU	Ok
35	P4807-05MS	BF-F14MS	BF140384.D	15 Nov 2024 00:58		RC/JU	Ok
36	P4807-05MSD	BF-F14MSD	BF140385.D	15 Nov 2024 01:24		RC/JU	Ok
37	P4807-01	TP-7	BF140386.D	15 Nov 2024 01:49		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111424

Review By	yogesh	Review On	11/15/2024 5:15:55 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:37 AM
SubDirectory	BF111424	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

38	P4798-01	MH-6	BF140387.D	15 Nov 2024 02:16	Need 2X Dilution	RC/JU	Dilution
39	P4789-01	BP-F21	BF140388.D	15 Nov 2024 02:42		RC/JU	Ok
40	P4793-01	M00-24-00345	BF140389.D	15 Nov 2024 03:08		RC/JU	Confirms

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111524

Review By	yogesh	Review On	11/18/2024 1:26:57 AM
Supervise By	mohammad	Supervise On	11/18/2024 2:51:39 AM
SubDirectory	BF111524	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140390.D	15 Nov 2024 09:16		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140391.D	15 Nov 2024 09:42		RC/JU	Ok,M
3	PB164789BL	PB164789BL	BF140392.D	15 Nov 2024 10:08		RC/JU	Ok
4	PB164789BS	PB164789BS	BF140393.D	15 Nov 2024 10:34		RC/JU	Ok,M
5	PB164886BS	PB164886BS	BF140394.D	15 Nov 2024 11:00		RC/JU	Ok,M
6	P4799-03	WC-TA2-02-C	BF140395.D	15 Nov 2024 11:33		RC/JU	Ok
7	P4799-03MS	WC-TA2-02-CMS	BF140396.D	15 Nov 2024 11:59		RC/JU	Ok,M
8	P4799-03MSD	WC-TA2-02-CMSD	BF140397.D	15 Nov 2024 12:25		RC/JU	Ok,M
9	P4799-07	WC-TA2-03-C	BF140398.D	15 Nov 2024 12:52		RC/JU	Ok
10	P4799-11	WC-TA1-01-C	BF140399.D	15 Nov 2024 13:18		RC/JU	Ok
11	P4799-15	WC-TA1-02-C	BF140400.D	15 Nov 2024 13:44		RC/JU	Ok
12	P4799-19	WC-TA1-03-C	BF140401.D	15 Nov 2024 14:10		RC/JU	Ok
13	P4807-04	TP-7	BF140402.D	15 Nov 2024 14:36		RC/JU	Ok
14	P4807-08	BF-F14	BF140403.D	15 Nov 2024 15:02		RC/JU	Ok
15	P4807-12	BF-F13	BF140404.D	15 Nov 2024 15:29		RC/JU	Ok
16	P4807-16	TP-6	BF140405.D	15 Nov 2024 15:55		RC/JU	Ok
17	P4809-04	MH-5	BF140406.D	15 Nov 2024 16:21		RC/JU	Ok
18	P4807-13	TP-6	BF140407.D	15 Nov 2024 16:47		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111524

Review By	yogesh	Review On	11/18/2024 1:26:57 AM		
Supervise By	mohammad	Supervise On	11/18/2024 2:51:39 AM		
SubDirectory	BF111524	HP Acquire Method	BNA_F	HP Processing Method	Bf111324
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12327,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4807-09	BF-F13	BF140408.D	15 Nov 2024 17:13		RC/JU	Ok
20	P4836-03	T3-STONE	BF140409.D	15 Nov 2024 17:40	Internal Standard Fail	RC/JU	ReRun
21	P4835-03	345-2-STONE	BF140410.D	15 Nov 2024 18:06		RC/JU	Ok
22	P4798-01DL	MH-6DL	BF140411.D	15 Nov 2024 18:32		RC/JU	Ok,M
23	P4825-01	STOCK-PILE	BF140412.D	15 Nov 2024 18:58		RC/JU	Ok,M
24	P4809-01	MH-5	BF140413.D	15 Nov 2024 19:25		RC/JU	Ok,M
25	P4850-01	OR-03-11142024	BF140414.D	15 Nov 2024 19:51		RC/JU	Ok,M
26	P4851-01	OK-02-11142024	BF140415.D	15 Nov 2024 20:17		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111624

Review By	yogesh	Review On	11/19/2024 3:19:51 AM
Supervise By	mohammad	Supervise On	11/19/2024 3:30:32 AM
SubDirectory	BF111624	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140416.D	16 Nov 2024 09:08		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140417.D	16 Nov 2024 09:34		RC/JU	Ok,M
3	PB164993BL	PB164993BL	BF140418.D	16 Nov 2024 10:02		RC/JU	Ok
4	PB164993BS	PB164993BS	BF140419.D	16 Nov 2024 10:28		RC/JU	Ok,M
5	P4821-04	BP-F-24	BF140420.D	16 Nov 2024 11:01	Internal Standard and Surrogate Fail	RC/JU	ReRun
6	P4821-08	BP-F-2	BF140421.D	16 Nov 2024 11:27		RC/JU	Ok
7	P4823-04	MH-4	BF140422.D	16 Nov 2024 11:53		RC/JU	Ok
8	P4833-04	MH-731	BF140423.D	16 Nov 2024 12:19		RC/JU	Ok
9	P4833-08	TP-2	BF140424.D	16 Nov 2024 12:46		RC/JU	Ok
10	P4824-01	T-20	BF140425.D	16 Nov 2024 13:12	Internal standard fail - hydrocarbons	RC/JU	Ok
11	P4821-05	BP-F-2	BF140426.D	16 Nov 2024 13:38		RC/JU	Ok
12	P4833-01	MH-731	BF140427.D	16 Nov 2024 14:04		RC/JU	Ok
13	P4833-01MS	MH-731MS	BF140428.D	16 Nov 2024 14:31		RC/JU	Ok,M
14	P4833-01MSD	MH-731MSD	BF140429.D	16 Nov 2024 14:57		RC/JU	Ok,M
15	P4848-01	TP-1	BF140430.D	16 Nov 2024 15:23		RC/JU	Ok,M
16	P4821-01	BP-F-24	BF140431.D	16 Nov 2024 15:49		RC/JU	Ok
17	P4821-01MS	BP-F-24MS	BF140432.D	16 Nov 2024 16:16		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111624

Review By	yogesh	Review On	11/19/2024 3:19:51 AM		
Supervise By	mohammad	Supervise On	11/19/2024 3:30:32 AM		
SubDirectory	BF111624	HP Acquire Method	BNA_F	HP Processing Method	Bf111324
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12327,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P4821-01MSD	BP-F-24MSD	BF140433.D	16 Nov 2024 16:42		RC/JU	Ok,M
19	P4823-01	MH-4	BF140434.D	16 Nov 2024 17:08		RC/JU	Ok,M
20	P4768-02	B-3-2	BF140435.D	16 Nov 2024 17:34		RC/JU	Ok,M
21	P4833-05	TP-2	BF140436.D	16 Nov 2024 18:00		RC/JU	Ok
22	P4849-01	RR-1	BF140437.D	16 Nov 2024 18:27	Internal Standard Fail	RC/JU	ReRun
23	P4768-04	B-4	BF140438.D	16 Nov 2024 18:53		RC/JU	Ok
24	P4768-03	B-3-3	BF140439.D	16 Nov 2024 19:19	Internal Standard Fail	RC/JU	ReRun
25	P4859-01	NB-08-111424	BF140440.D	16 Nov 2024 19:45	Internal Standard Fail	RC/JU	ReRun
26	P4857-01	ETGI-328	BF140441.D	16 Nov 2024 20:11	Internal standard fail - hydrocarbons and oily matrix	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111924

Review By	yogesh	Review On	11/20/2024 7:05:01 AM
Supervise By	mohammad	Supervise On	11/20/2024 7:31:21 AM
SubDirectory	BF111924	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140462.D	19 Nov 2024 09:32		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140463.D	19 Nov 2024 10:34		RC/JU	Ok,M
3	PB165019TB	PB165019TB	BF140464.D	19 Nov 2024 11:00		RC/JU	Ok
4	P4848-02MS	TP-1MS	BF140465.D	19 Nov 2024 11:36		RC/JU	Ok,M
5	P4848-02MSD	TP-1MSD	BF140466.D	19 Nov 2024 12:02		RC/JU	Ok,M
6	P4870-04	MH-735	BF140467.D	19 Nov 2024 12:28		RC/JU	Ok
7	P4852-01	COMP-1	BF140468.D	19 Nov 2024 12:54		RC/JU	Ok,M
8	P4852-02	COMP-2	BF140469.D	19 Nov 2024 13:21		RC/JU	Ok,M
9	P4852-03	COMP-3	BF140470.D	19 Nov 2024 13:47		RC/JU	Ok,M
10	P4852-03MS	COMP-3MS	BF140471.D	19 Nov 2024 14:13		RC/JU	Ok,M
11	P4852-03MSD	COMP-3MSD	BF140472.D	19 Nov 2024 14:39		RC/JU	Ok,M
12	P4771-05	P-1	BF140473.D	19 Nov 2024 15:06		RC/JU	Ok,M
13	P4771-07	P-2	BF140474.D	19 Nov 2024 15:32		RC/JU	Ok,M
14	P4889-03	#72-11978	BF140475.D	19 Nov 2024 15:58		RC/JU	Ok,M
15	P4869-01	EO-02-11152024	BF140476.D	19 Nov 2024 16:24		RC/JU	Ok
16	P4870-01	TP-1	BF140477.D	19 Nov 2024 16:51		RC/JU	Ok
17	P4857-03	VNJ-226	BF140478.D	19 Nov 2024 17:17		RC/JU	Ok,M
18	P4889-05	#72-11930	BF140479.D	19 Nov 2024 17:43	Need 5X Dilution	RC/JU	Dilution

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111924

Review By	yogesh	Review On	11/20/2024 7:05:01 AM		
Supervise By	mohammad	Supervise On	11/20/2024 7:31:21 AM		
SubDirectory	BF111924	HP Acquire Method	BNA_F	HP Processing Method	Bf111324
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12327,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4889-01	VNJ-228	BF140480.D	19 Nov 2024 18:10		RC/JU	Ok,M
20	P4849-01	RR-1	BF140481.D	19 Nov 2024 18:36		RC/JU	Ok,M
21	P4768-03	B-3-3	BF140482.D	19 Nov 2024 19:02		RC/JU	Ok
22	P4859-01RE	NB-08-111424RE	BF140483.D	19 Nov 2024 19:28	Fax is already given	RC/JU	Confirms
23	P4908-01	SP-1	BF140484.D	19 Nov 2024 19:54		RC/JU	Ok
24	P4908-03	SP-2	BF140485.D	19 Nov 2024 20:20		RC/JU	Ok,M
25	P4768-07	B-6-3	BF140486.D	19 Nov 2024 20:47		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112024

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF112024	HP Acquire Method	BNA_F	HP Processing Method	Bf111324
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12329,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140487.D	20 Nov 2024 09:04		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140488.D	20 Nov 2024 09:31		RC/JU	Ok,NR
3	PB165039BL	PB165039BL	BF140489.D	20 Nov 2024 09:57		RC/JU	Ok
4	PB165039BS	PB165039BS	BF140490.D	20 Nov 2024 10:23		RC/JU	Ok,NR
5	PB165039BSD	PB165039BSD	BF140491.D	20 Nov 2024 10:49		RC/JU	Ok,NR
6	P4843-01	SW-WTS-01	BF140492.D	20 Nov 2024 11:23		RC/JU	Ok,NR
7	P4868-01	RW5-SP100-20241114	BF140493.D	20 Nov 2024 11:49		RC/JU	Ok
8	P4868-03	RW5-SP303-20241114	BF140494.D	20 Nov 2024 12:15	Internal Standard Fail	RC/JU	ReRun
9	P4892-02	WB-310-BOT	BF140495.D	20 Nov 2024 12:42		RC/JU	Ok
10	P4892-02MS	WB-310-BOTMS	BF140496.D	20 Nov 2024 13:08		RC/JU	Ok,NR
11	P4892-02MSD	WB-310-BOTMSD	BF140497.D	20 Nov 2024 13:35		RC/JU	Ok,NR
12	P4868-03RE	RW5-SP303-20241114	BF140498.D	20 Nov 2024 14:01	Internal Standard Fail	RC/JU	Confirms
13	SSTDCCC040	SSTDCCC040EC	BF140499.D	20 Nov 2024 14:34		RC/JU	Ok,NR
14	DFTPP	DFTPP	BF140500.D	20 Nov 2024 15:33		RC/JU	Ok
15	SSTDCCC040	SSTDCCC040	BF140501.D	20 Nov 2024 15:59		RC/JU	Ok,NR
16	PB165086BL	PB165086BL	BF140502.D	20 Nov 2024 16:25		RC/JU	Ok
17	P4849-02	RR-1	BF140503.D	20 Nov 2024 16:58		RC/JU	Ok
18	P4870-07	MH-736	BF140504.D	20 Nov 2024 17:24		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112024

Review By	Review On
Supervise By	Supervise On
SubDirectory	BF112024
HP Acquire Method	BNA_F
HP Processing Method	Bf111324
STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12329,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4910-01	MH-COTTAGE	BF140505.D	20 Nov 2024 17:51		RC/JU	Ok
20	P4893-01	MH-763	BF140506.D	20 Nov 2024 18:17		RC/JU	Ok
21	P4893-05	MH-762	BF140507.D	20 Nov 2024 18:43		RC/JU	Ok
22	P4768-05	B-5	BF140508.D	20 Nov 2024 19:09	Internal Standard Fail	RC/JU	Ok
23	P4870-10	TP-15	BF140509.D	20 Nov 2024 19:35		RC/JU	Ok
24	P4892-01	WB-310-TOP	BF140510.D	20 Nov 2024 20:02		RC/JU	Ok,NR
25	P4909-01	BU-02-111824	BF140511.D	20 Nov 2024 20:28		RC/JU	Ok,NR
26	P4910-05	MH-759	BF140512.D	20 Nov 2024 20:54	Internal Standard Fail	RC/JU	Ok,NR
27	P4889-05DL	#72-11930DL	BF140513.D	20 Nov 2024 21:21	Internal Standard Fail	RC/JU	Ok,NR
28	P4822-10	SOIL-5	BF140514.D	20 Nov 2024 21:47		RC/JU	Ok,NR
29	P4822-12	SOIL-6	BF140515.D	20 Nov 2024 22:13		RC/JU	Ok,NR
30	P4822-06	SOIL-3	BF140516.D	20 Nov 2024 22:40		RC/JU	Ok,NR
31	P4822-08	SOIL-4	BF140517.D	20 Nov 2024 23:06		RC/JU	Ok,NR
32	P4822-02	SOIL-1	BF140518.D	20 Nov 2024 23:32	Internal Standard Fail	RC/JU	ReRun
33	P4822-04	SOIL-2	BF140519.D	20 Nov 2024 23:58	Internal Standard Fail	RC/JU	ReRun
34	P4887-05	MH-760	BF140520.D	21 Nov 2024 00:24	Internal Standard Fail	RC/JU	ReRun
35	P4860-02	PH2-BOT-001	BF140521.D	21 Nov 2024 00:51	Internal Standard Fail	RC/JU	ReRun
36	P4860-03	PH2-BOT-002	BF140522.D	21 Nov 2024 01:17	Internal Standard Fail	RC/JU	ReRun
37	P4860-07	PH2-BOT-008	BF140523.D	21 Nov 2024 01:43	Internal Standard Fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112024

Review By		Review On	
Supervise By		Supervise On	
SubDirectory	BF112024	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC	SP6624		
Internal Standard/PEM	S12329,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	P4916-01	TP-1-WC	BF140524.D	21 Nov 2024 02:09	Internal Standard Fail	RC/JU	ReRun
39	P4916-05	TP-2-WC	BF140525.D	21 Nov 2024 02:35	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration