

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF111524

SequenceID : BF111524

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
5. GC/MS Calibration Met:	_____	_____	_____✓_____
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓_____	_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

____ ✓ ____

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

____ ✓ ____

a. B/N Fraction Recovery and RPD failed for some compound in P4799-03MS/MSD due to matrix interference. No corrective action is required.

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

____ ✓ ____

Internal standard failed in the sample P4836-03 due to the presence of non-targeted hydrocarbons, which can be witnessed by the abnormal chromatogram. Hence no corrective action is needed.

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

____ ✓ ____

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

____ ✓ ____

ADDITIONAL COMMENTS:

Compounds #69 and #9 being poor recovery compounds their recovery was low in PB164886BS. The data will be used for hardcopies. The samples P4825-01, P4809-01, P4850-01, P4851-01 had to be analyzed with 10X dilution each due to dirty and viscous matrix, which was difficult to inject otherwise. Hence no corrective action is required.

krunal

Analyst

11/16/2024

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BF111524

SequenceID : BF111524 NA NO YES