

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/08/2024

Lab File ID: BF140286.D Time Analyzed: 09:35

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	203786	6.881	760595	8.16	473134	9.92
UPPER LIMIT	407572	7.381	1521190	8.663	946268	10.422
LOWER LIMIT	101893	6.381	380298	7.663	236567	9.422
EPA SAMPLE NO.						
01 BP-B4MS	111731	6.88	394140	8.16	192644 *	9.92
02 BP-B4MSD	110036	6.88	391943	8.16	195155 *	9.92
03 B-3-1	113687	6.88	388933	8.16	199747 *	9.92
04 B-6-2	108852	6.88	376967 *	8.16	191196 *	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.