

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/15/2024

Lab File ID: BF140391.D Time Analyzed: 09:42

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	137586	6.881	510501	8.16	282269	9.92
UPPER LIMIT	275172	7.381	1021000	8.657	564538	10.416
LOWER LIMIT	68793	6.381	255251	7.657	141135	9.416
EPA SAMPLE NO.						
01 PB164789BL	139459	6.88	521031	8.16	292158	9.91
02 PB164789BS	137404	6.88	514072	8.16	287836	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.