

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/16/2024

Lab File ID: BF140417.D Time Analyzed: 09:34

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	171491	6.875	639592	8.16	364979	9.92
UPPER LIMIT	342982	7.375	1279180	8.657	729958	10.416
LOWER LIMIT	85745.5	6.375	319796	7.657	182490	9.416
EPA SAMPLE NO.						
01 B-3-2	116979	6.88	433268	8.15	227549	9.91
02 B-4	126758	6.88	464845	8.15	232288	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.