

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/19/2024

Lab File ID: BF140463.D Time Analyzed: 10:34

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	524892	11.398	344697	14.069	292973	15.592
UPPER LIMIT	1049780	11.898	689394	14.569	585946	16.092
LOWER LIMIT	262446	10.898	172349	13.569	146487	15.092
EPA SAMPLE NO.						
01 B-3-3	406415	11.40	265711	14.05	205466	15.53
02 B-6-3	337964	11.40	204768	14.05	163923	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.