

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JPCL01

Lab Code: CHEM

SAS No.: P4768 SDG NO.: P4768

Lab File ID: BF140500.D

DFTPP Injection Date: 11/20/2024

Instrument ID: BNA_F

DFTPP Injection Time: 15:33

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	37.7
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	38.4
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	50
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	28.1
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	14.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.9 (18.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140501.D	11/20/2024	15:59
B-5	P4768-05	BF140508.D	11/20/2024	19:09