

Surrogate Summary

SDG No.: **P4768**

Client: **JPCL Engineering**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4768-01	B-3-1	1,2-Dichloroethane-d4	50	56.3	113	50	163
		Dibromofluoromethane	50	51.4	103	54	147
		Toluene-d8	50	49.9	100	58	134
		4-Bromofluorobenzene	50	48.5	97	29	146
P4768-02	B-3-2	1,2-Dichloroethane-d4	50	56.5	113	50	163
		Dibromofluoromethane	50	52.4	105	54	147
		Toluene-d8	50	49.8	100	58	134
		4-Bromofluorobenzene	50	47.9	96	29	146
P4768-03	B-3-3	1,2-Dichloroethane-d4	50	60.1	120	50	163
		Dibromofluoromethane	50	52.9	106	54	147
		Toluene-d8	50	49.7	99	58	134
		4-Bromofluorobenzene	50	48.5	97	29	146
P4768-04	B-4	1,2-Dichloroethane-d4	50	34.0	68	50	163
		Dibromofluoromethane	50	48.9	98	54	147
		Toluene-d8	50	41.5	83	58	134
		4-Bromofluorobenzene	50	23.0	46	29	146
P4768-05	B-5	1,2-Dichloroethane-d4	50	61.5	123	50	163
		Dibromofluoromethane	50	52.7	105	54	147
		Toluene-d8	50	51.2	102	58	134
		4-Bromofluorobenzene	50	57.6	115	29	146
P4768-06	B-6-2	1,2-Dichloroethane-d4	50	57.0	114	50	163
		Dibromofluoromethane	50	52.2	104	54	147
		Toluene-d8	50	48.9	98	58	134
		4-Bromofluorobenzene	50	46.0	92	29	146
P4768-07	B-6-3	1,2-Dichloroethane-d4	50	60.8	122	50	163
		Dibromofluoromethane	50	52.3	105	54	147
		Toluene-d8	50	50.5	101	58	134
		4-Bromofluorobenzene	50	50.8	102	29	146
VY1111SBL01	VY1111SBL01	1,2-Dichloroethane-d4	50	43.3	87	50	163
		Dibromofluoromethane	50	46.4	93	54	147
		Toluene-d8	50	47.9	96	58	134
		4-Bromofluorobenzene	50	40.0	80	29	146
VY1111SBS01	VY1111SBS01	1,2-Dichloroethane-d4	50	52.3	105	50	163
		Dibromofluoromethane	50	53.4	107	54	147
		Toluene-d8	50	52.1	104	58	134
		4-Bromofluorobenzene	50	53.2	106	29	146
VY1111SBSD01	VY1111SBSD01	1,2-Dichloroethane-d4	50	53.5	107	50	163
		Dibromofluoromethane	50	52.1	104	54	147
		Toluene-d8	50	50.3	100	58	134
		4-Bromofluorobenzene	50	52.0	104	29	146