

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JPCL01
 Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768
 Lab File ID: VY020239.D Date Analyzed: 11/11/2024
 Instrument ID: MSVOA_Y Time Analyzed: 11:11
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	187296	13.346				
UPPER LIMIT	374592	13.846				
LOWER LIMIT	93648	12.846				
EPA SAMPLE NO.						
B-3-1	133835	13.35				
B-3-2	117809	13.35				
B-3-3	129893	13.35				
B-4	3939 *	13.35				
B-5	144633	13.35				
B-6-2	128461	13.35				
B-6-3	131499	13.35				
VY1111SBL01	111959	13.35				
VY1111SBS01	179570	13.35				
VY1111SBSD01	176083	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.