

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JPCL01  
 Lab Code: CHEM Case No.: P4768 SAS No.: P4768 SDG NO.: P4768  
 Lab File ID: VY020347.D Date Analyzed: 11/20/2024  
 Instrument ID: MSVOA\_Y Time Analyzed: 09:34  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 139421        | 13.352 |  |  |  |  |
| UPPER LIMIT    | 278842        | 13.852 |  |  |  |  |
| LOWER LIMIT    | 69710.5       | 12.852 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| B-4            | 98434         | 13.35  |  |  |  |  |
| VY1120SBL01    | 104132        | 13.35  |  |  |  |  |
| VY1120SBS01    | 135305        | 13.35  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.