

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: M2AS01
Lab Code: CHEM Case No.: P4770 SAS No.: P4770 SDG NO.: P4770
Lab File ID: VN084960.D Date Analyzed: 11/20/2024
Instrument ID: MSVOA_N Time Analyzed: 11:18
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	164992	8.22	271316	9.10	240365	11.87
UPPER LIMIT	329984	8.724	542632	9.6	480730	12.365
LOWER LIMIT	82496	7.724	135658	8.6	120183	11.365
EPA SAMPLE NO.						
MW-1	152380	8.22	277071	9.10	242199	11.87
VN1120WBL01	178908	8.22	319489	9.10	277659	11.87
VN1120WBS02	150630	8.22	253557	9.10	229692	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.