

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: M2AS01
 Lab Code: CHEM Case No.: P4770 SAS No.: P4770 SDG NO.: P4770
 Lab File ID: VN084935.D Date Analyzed: 11/19/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:27
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	140060	13.788				
UPPER LIMIT	280120	14.288				
LOWER LIMIT	70030	13.288				
EPA SAMPLE NO.						
MW-3	148014	13.79				
VN1119WBL01	128229	13.79				
VN1119WBS01	131165	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.