

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: M2AS01
 Lab Code: CHEM Case No.: P4770 SAS No.: P4770 SDG NO.: P4770
 Lab File ID: VN084960.D Date Analyzed: 11/20/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:18
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	122785	13.788				
UPPER LIMIT	245570	14.288				
LOWER LIMIT	61392.5	13.288				
EPA SAMPLE NO.						
MW-1	122080	13.79				
VN1120WBL01	123641	13.79				
VN1120WBS02	114872	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.