

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: M2AS01
Lab Code: CHEM Case No.: P4770 SAS No.: P4770 SDG NO.: P4770
Lab File ID: VN084814.D BFB Injection Date: 11/13/2024
Instrument ID: MSVOA_N BFB Injection Time: 08:53
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.2
75	30.0 - 60.0% of mass 95	50.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.7 (0.9) 1
174	50.0 - 100.0% of mass 95	76.3
175	5.0 - 9.0% of mass 174	5.7 (7.5) 1
176	95.0 - 101.0% of mass 174	75.8 (99.3) 1
177	5.0 - 9.0% of mass 176	5 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN084815.D	11/13/2024	09:52
VN1113WBL01	VN1113WBL01	VN084817.D	11/13/2024	10:50
VN1113WBS01	VN1113WBS01	VN084822.D	11/13/2024	13:24
VN1113WBSD01	VN1113WBSD01	VN084823.D	11/13/2024	14:00
FIELD-BLANK	P4770-04	VN084830.D	11/13/2024	16:47
TRIP-BLANK	P4770-05	VN084832.D	11/13/2024	17:35