

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4774 SAS No.: P4774 SDG NO.: P4774

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 11/14/2024

Lab File ID: BN035082.D Time Analyzed: 10:12

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2551	7.553	6362	10.32	5037	14.19
UPPER LIMIT	5102	8.053	12724	10.819	10074	14.69
LOWER LIMIT	1275.5	7.053	3181	9.819	2518.5	13.69
EPA SAMPLE NO.						
01 VPB190-HYD-20241106RE	2635	7.55	6903	10.32	5180	14.19

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.