

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4774 SAS No.: P4774 SDG No.: P4774

Instrument ID: MSVOA\_N Calibration Date/Time: 11/08/2024 09:39

Lab File ID: VN084737.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Chloromethane                  | 0.952 | 0.557  | 0.1        | -41.49 | 20    |
| Vinyl Chloride                 | 0.618 | 0.547  |            | -11.49 | 20    |
| Bromomethane                   | 0.328 | 0.287  |            | -12.5  | 20    |
| Chloroethane                   | 0.488 | 0.357  |            | -26.84 | 20    |
| Trichlorofluoromethane         | 1.018 | 0.949  |            | -6.78  | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.575 | 0.549  |            | -4.52  | 20    |
| 1,1-Dichloroethene             | 0.569 | 0.508  |            | -10.72 | 20    |
| Acetone                        | 0.238 | 0.238  |            | 0      | 20    |
| Carbon Disulfide               | 1.753 | 1.371  |            | -21.79 | 20    |
| Methyl tert-butyl Ether        | 1.745 | 1.787  |            | 2.41   | 20    |
| Methylene Chloride             | 0.635 | 0.580  |            | -8.66  | 20    |
| trans-1,2-Dichloroethene       | 0.585 | 0.522  |            | -10.77 | 20    |
| 1,1-Dichloroethane             | 1.102 | 1.027  | 0.1        | -6.81  | 20    |
| 2-Butanone                     | 0.337 | 0.361  |            | 7.12   | 20    |
| Carbon Tetrachloride           | 0.525 | 0.529  |            | 0.76   | 20    |
| cis-1,2-Dichloroethene         | 0.683 | 0.647  |            | -5.27  | 20    |
| Chloroform                     | 1.121 | 1.095  |            | -2.32  | 20    |
| 1,1,1-Trichloroethane          | 1.021 | 0.989  |            | -3.13  | 20    |
| Methylcyclohexane              | 0.478 | 0.502  |            | 5.02   | 20    |
| Benzene                        | 1.507 | 1.426  |            | -5.38  | 20    |
| 1,2-Dichloroethane             | 0.488 | 0.484  |            | -0.82  | 20    |
| Trichloroethene                | 0.348 | 0.335  |            | -3.74  | 20    |
| 1,2-Dichloropropane            | 0.354 | 0.349  |            | -1.41  | 20    |
| Bromodichloromethane           | 0.526 | 0.525  |            | -0.19  | 20    |
| 4-Methyl-2-Pentanone           | 0.406 | 0.443  |            | 9.11   | 20    |
| Toluene                        | 0.882 | 0.897  |            | 1.7    | 20    |
| t-1,3-Dichloropropene          | 0.544 | 0.542  |            | -0.37  | 20    |
| cis-1,3-Dichloropropene        | 0.576 | 0.582  |            | 1.04   | 20    |
| 1,1,2-Trichloroethane          | 0.332 | 0.338  |            | 1.81   | 20    |
| 2-Hexanone                     | 0.296 | 0.342  |            | 15.54  | 20    |
| Dibromochloromethane           | 0.378 | 0.407  |            | 7.67   | 20    |
| Tetrachloroethene              | 0.333 | 0.341  |            | 2.4    | 20    |
| Chlorobenzene                  | 1.119 | 1.084  | 0.3        | -3.13  | 20    |
| Ethyl Benzene                  | 1.867 | 1.938  |            | 3.8    | 20    |
| m/p-Xylenes                    | 0.707 | 0.733  |            | 3.68   | 20    |
| o-Xylene                       | 0.661 | 0.705  |            | 6.66   | 20    |
| Styrene                        | 1.154 | 1.220  |            | 5.72   | 20    |
| Bromoform                      | 0.293 | 0.314  | 0.1        | 7.17   | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4774 SAS No.: P4774 SDG No.: P4774

Instrument ID: MSVOA\_N Calibration Date/Time: 11/08/2024 09:39

Lab File ID: VN084737.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                  | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|---------------------------|-------|--------|------------|--------|-------|
| Isopropylbenzene          | 3.504 | 3.511  |            | 0.2    | 20    |
| 1,1,2,2-Tetrachloroethane | 1.170 | 1.092  | 0.3        | -6.67  | 20    |
| 1,3-Dichlorobenzene       | 1.838 | 1.648  |            | -10.34 | 20    |
| 1,4-Dichlorobenzene       | 1.937 | 1.652  |            | -14.71 | 20    |
| 1,2-Dichlorobenzene       | 1.766 | 1.642  |            | -7.02  | 20    |
| 1,2-Dichloroethane-d4     | 0.722 | 0.672  |            | -6.93  | 20    |
| Dibromofluoromethane      | 0.338 | 0.337  |            | -0.3   | 20    |
| Toluene-d8                | 1.247 | 1.241  |            | -0.48  | 20    |
| 4-Bromofluorobenzene      | 0.466 | 0.466  |            | 0      | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.