

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
 Data File : VN084787.D
 Acq On : 11 Nov 2024 21:18
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Nov 12 00:39:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	176515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	291331	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	261381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	128017	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118955	46.659	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.320%
35) Dibromofluoromethane	8.165	113	97314	49.349	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	10.565	98	351476	48.386	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.780%
62) 4-Bromofluorobenzene	12.847	95	135180	49.793	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.580%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.118	85	84501	41.643	ug/l	96
3) Chloromethane	2.360	50	90304	37.285	ug/l	96
4) Vinyl Chloride	2.512	62	99906	45.776	ug/l	97
5) Bromomethane	2.901	94	51624	44.644	ug/l	99
6) Chloroethane	3.054	64	65003	48.061	ug/l	99
7) Trichlorofluoromethane	3.454	101	163963	45.606	ug/l	99
8) Diethyl Ether	3.948	74	59294	46.753	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.348	101	92258	45.414	ug/l	97
10) Methyl Iodide	4.571	142	120168	44.260	ug/l	98
11) Tert butyl alcohol	5.548	59	106297	315.868	ug/l	99
12) 1,1-Dichloroethene	4.318	96	85753	42.660	ug/l	98
13) Acrolein	4.171	56	83096	247.625	ug/l	99
14) Allyl chloride	5.006	41	144576	44.349	ug/l	91
15) Acrylonitrile	5.718	53	245495	251.991	ug/l	99
16) Acetone	4.430	43	198198	269.348	ug/l	93
17) Carbon Disulfide	4.689	76	223087	36.038	ug/l	99
18) Methyl Acetate	5.018	43	130899	48.542	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	305562	49.595	ug/l	96
20) Methylene Chloride	5.259	84	98051	43.728	ug/l	86
21) trans-1,2-Dichloroethene	5.771	96	88013	42.641	ug/l	97
22) Diisopropyl ether	6.671	45	298223	46.040	ug/l	96
23) Vinyl Acetate	6.595	43	1081205	242.044	ug/l	98
24) 1,1-Dichloroethane	6.565	63	173686	44.661	ug/l	99
25) 2-Butanone	7.483	43	325803	273.920	ug/l	98
26) 2,2-Dichloropropane	7.483	77	140167	41.411	ug/l	98
27) cis-1,2-Dichloroethene	7.483	96	108592	45.057	ug/l	96
28) Bromochloromethane	7.806	49	72979	47.101	ug/l	92
29) Tetrahydrofuran	7.836	42	205582	260.536	ug/l	94
30) Chloroform	7.959	83	185103	46.764	ug/l	99
31) Cyclohexane	8.253	56	141733	40.267	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	168045	46.644	ug/l	98
36) 1,1-Dichloropropene	8.365	75	119608	43.735	ug/l	97
37) Ethyl Acetate	7.553	43	123291	49.688	ug/l	97
38) Carbon Tetrachloride	8.353	117	147502	48.253	ug/l	97
39) Methylcyclohexane	9.594	83	133862	48.113	ug/l	99
40) Benzene	8.600	78	391445	44.569	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
 Data File : VN084787.D
 Acq On : 11 Nov 2024 21:18
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Nov 12 00:39:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	68548	51.372	ug/l	95
42) 1,2-Dichloroethane	8.665	62	133491	46.931	ug/l	99
43) Isopropyl Acetate	8.688	43	243452	53.170	ug/l	99
44) Trichloroethene	9.347	130	91914	45.377	ug/l	100
45) 1,2-Dichloropropane	9.618	63	95605	46.403	ug/l	95
46) Dibromomethane	9.706	93	67476	46.413	ug/l	96
47) Bromodichloromethane	9.888	83	145902	47.610	ug/l #	97
48) Methyl methacrylate	9.677	41	100305	51.941	ug/l	96
49) 1,4-Dioxane	9.694	88	38986	1139.238	ug/l #	80
51) 4-Methyl-2-Pentanone	10.447	43	641767	271.309	ug/l	98
52) Toluene	10.630	92	246524	47.993	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	146707	46.248	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	153802	45.810	ug/l	98
55) 1,1,2-Trichloroethane	11.012	97	93404	48.273	ug/l	95
56) Ethyl methacrylate	10.871	69	156659	55.974	ug/l	95
57) 1,3-Dichloropropane	11.159	76	160452	48.318	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	352227	268.948	ug/l	97
59) 2-Hexanone	11.194	43	486706	282.678	ug/l	97
60) Dibromochloromethane	11.359	129	111721	50.662	ug/l	99
61) 1,2-Dibromoethane	11.465	107	96789	48.848	ug/l	99
64) Tetrachloroethene	11.100	164	79480	45.715	ug/l	99
65) Chlorobenzene	11.888	112	263506	45.063	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	95834	47.117	ug/l	99
67) Ethyl Benzene	11.959	91	476489	48.816	ug/l	100
68) m/p-Xylenes	12.071	106	359947	97.406	ug/l	99
69) o-Xylene	12.394	106	174747	50.550	ug/l	98
70) Styrene	12.412	104	296821	49.221	ug/l	99
71) Bromoform	12.576	173	76363	49.894	ug/l #	100
73) Isopropylbenzene	12.694	105	461999	51.499	ug/l	100
74) N-amyl acetate	12.494	43	198094	51.881	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.935	83	136784	45.679	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	178471	69.743	ug/l	57
77) Bromobenzene	12.976	156	106796	44.282	ug/l	95
78) n-propylbenzene	13.035	91	513794	48.874	ug/l	97
79) 2-Chlorotoluene	13.124	91	317997	46.294	ug/l	98
80) 1,3,5-Trimethylbenzene	13.171	105	377354	50.751	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	54154	50.205	ug/l	90
82) 4-Chlorotoluene	13.218	91	317522	43.924	ug/l	97
83) tert-Butylbenzene	13.435	119	333151	54.138	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	381057	49.655	ug/l	99
85) sec-Butylbenzene	13.618	105	440437	50.669	ug/l	100
86) p-Isopropyltoluene	13.729	119	374128	52.480	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	195927	41.628	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	197206	45.560	ug/l	99
89) n-Butylbenzene	14.053	91	310726	46.596	ug/l	98
90) Hexachloroethane	14.329	117	69240	43.552	ug/l	93
91) 1,2-Dichlorobenzene	14.106	146	203390	44.970	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.718	75	31583	52.063	ug/l	93
93) 1,2,4-Trichlorobenzene	15.394	180	117324	47.394	ug/l	99
94) Hexachlorobutadiene	15.500	225	48175	44.315	ug/l	100
95) Naphthalene	15.641	128	402741	54.353	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	115194	47.936	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111124\
Data File : VN084787.D
Acq On : 11 Nov 2024 21:18
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Nov 12 00:39:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 176515
Ion Ratio Lower Upper
168 100
99 61.9 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 41.643 ug/l
RT: 2.118 min Scan# 28
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 85 Resp: 84501
Ion Ratio Lower Upper
85 100
87 32.9 17.6 52.9









#3
Chloromethane
Concen: 37.285 ug/l
RT: 2.360 min Scan# 61
Delta R.T. 0.001 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 90304
Ion Ratio Lower Upper
50 100
52 31.2 26.8 40.2



#4
Vinyl Chloride
Concen: 45.776 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 62 Resp: 99906
Ion Ratio Lower Upper
62 100
64 31.1 26.2 39.4









#5
Bromomethane
Concen: 44.644 ug/l
RT: 2.901 min Scan# 11
Delta R.T. -0.052 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 51624
Ion Ratio Lower Upper
94 100
96 90.0 72.9 109.3



#6
Chloroethane
Concen: 48.061 ug/l
RT: 3.054 min Scan# 187
Delta R.T. -0.064 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 64 Resp: 65003
Ion Ratio Lower Upper
64 100
66 29.8 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 45.606 ug/l
 RT: 3.454 min Scan# 21
 Delta R.T. -0.041 min
 Lab File: VN084787.D
 Acq: 11 Nov 2024 21:18




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 163963
 Ion Ratio Lower Upper
 101 100
 103 64.0 51.8 77.6



#8
 Diethyl Ether
 Concen: 46.753 ug/l
 RT: 3.948 min Scan# 339
 Delta R.T. -0.011 min
 Lab File: VN084787.D
 Acq: 11 Nov 2024 21:18

Tgt Ion: 74 Resp: 59294
 Ion Ratio Lower Upper
 74 100
 45 91.7 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.414 ug/l
RT: 4.348 min Scan# 407
Delta R.T. -0.023 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 92258
Ion Ratio Lower Upper
101 100
85 44.7 36.2 54.4
151 77.1 59.3 88.9



#10
Methyl Iodide
Concen: 44.260 ug/l
RT: 4.571 min Scan# 445
Delta R.T. -0.023 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion:142 Resp: 120168
Ion Ratio Lower Upper
142 100
127 44.7 35.3 52.9
141 15.4 11.0 16.6









#11
Tert butyl alcohol
Concen: 315.868 ug/l
RT: 5.548 min Scan# 611
Delta R.T. 0.024 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 106297
Ion Ratio Lower Upper
59 100
57 9.7 8.2 12.2



#12
1,1-Dichloroethene
Concen: 42.660 ug/l
RT: 4.318 min Scan# 402
Delta R.T. -0.023 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18


Tgt Ion: 96 Resp: 85753
Ion Ratio Lower Upper
96 100
61 163.9 128.6 192.8
98 65.3 51.7 77.5








#13
Acrolein
Concen: 247.625 ug/l
RT: 4.171 min Scan# 31
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC







Tgt Ion: 56 Resp: 83096
Ion Ratio Lower Upper
56 100
55 70.4 57.2 85.8



#14
Allyl chloride
Concen: 44.349 ug/l
RT: 5.006 min Scan# 519
Delta R.T. -0.018 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 41 Resp: 144576
Ion Ratio Lower Upper
41 100
39 64.4 59.0 88.6
76 34.3 29.8 44.6









#15
Acrylonitrile
Concen: 251.991 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 245495
Ion Ratio Lower Upper
53 100
52 81.8 66.2 99.4
51 35.4 28.8 43.2



#16
Acetone
Concen: 269.348 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 43 Resp: 198198
Ion Ratio Lower Upper
43 100
58 34.2 24.4 36.6





#17

Carbon Disulfide

Concen: 36.038 ug/l

RT: 4.689 min Scan# 4

Delta R.T. -0.023 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 223087

Ion Ratio Lower Upper

76 100

78 9.3 7.3 10.9



#18

Methyl Acetate

Concen: 48.542 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN084787.D




Acq: 11 Nov 2024 21:18


Tgt Ion: 43 Resp: 130899

Ion Ratio Lower Upper

43 100

74 24.2 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 49.595 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 305562
Ion Ratio Lower Upper
73 100
57 20.2 17.7 26.5



#20
Methylene Chloride
Concen: 43.728 ug/l
RT: 5.259 min Scan# 562
Delta R.T. -0.018 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 84 Resp: 98051
Ion Ratio Lower Upper
84 100
49 112.9 107.2 160.8
51 34.3 31.8 47.8
86 57.8 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 42.641 ug/l
RT: 5.771 min Scan# 649
Delta R.T. -0.018 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 88013
Ion Ratio Lower Upper
96 100
61 131.0 108.4 162.6
98 61.6 49.8 74.8



#22
Diisopropyl ether
Concen: 46.040 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 45 Resp: 298223
Ion Ratio Lower Upper
45 100
43 49.5 42.3 63.5
87 29.4 21.8 32.6
59 12.3 9.6 14.4









#23
 Vinyl Acetate
 Concen: 242.044 ug/l
 RT: 6.595 min Scan# 71
 Delta R.T. -0.006 min
 Lab File: VN084787.D
 Acq: 11 Nov 2024 21:18




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1081205
 Ion Ratio Lower Upper
 43 100
 86 11.3 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 44.661 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.006 min
 Lab File: VN084787.D
 Acq: 11 Nov 2024 21:18

Tgt Ion: 63 Resp: 173686
 Ion Ratio Lower Upper
 63 100
 98 6.6 3.4 10.2
 100 3.7 2.0 5.8









#25
2-Butanone
Concen: 273.920 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 325803
Ion Ratio Lower Upper
43 100
72 27.7 21.4 32.2



#26
2,2-Dichloropropane
Concen: 41.411 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 77 Resp: 140167
Ion Ratio Lower Upper
77 100
97 22.5 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 45.057 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. -0.006 min
 Lab File: VN084787.D
 Acq: 11 Nov 2024 21:18

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 96 Resp: 108592




Ion	Ratio	Lower	Upper
96	100		
61	143.9	0.0	301.8
98	64.5	0.0	128.2




#28
 Bromochloromethane
 Concen: 47.101 ug/l
 RT: 7.806 min Scan# 995
 Delta R.T. -0.006 min
 Lab File: VN084787.D
 Acq: 11 Nov 2024 21:18

Tgt Ion: 49 Resp: 72979

Ion	Ratio	Lower	Upper
49	100		
129	1.4	0.0	3.6
130	79.7	58.0	87.0





#29

Tetrahydrofuran

Concen: 260.536 ug/l

RT: 7.836 min Scan# 1000

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 205582

Ion Ratio Lower Upper

42 100

72 47.1 34.6 51.8

71 43.5 32.2 48.4



#30

Chloroform

Concen: 46.764 ug/l

RT: 7.959 min Scan# 1021

Delta R.T. -0.006 min

Lab File: VN084787.D




Acq: 11 Nov 2024 21:18


Tgt Ion: 83 Resp: 185103

Ion Ratio Lower Upper

83 100

85 64.2 51.0 76.4









#31
Cyclohexane
Concen: 40.267 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 141733
Ion Ratio Lower Upper
56 100
69 31.2 23.8 35.6
84 87.1 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 46.644 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 97 Resp: 168045
Ion Ratio Lower Upper
97 100
99 62.9 51.2 76.8
61 44.8 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 46.659 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 118955

Ion Ratio Lower Upper

65 100

67 50.9 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.094 min Scan# 1214

Delta R.T. -0.006 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18




Tgt Ion: 114 Resp: 291331


Ion Ratio Lower Upper

114 100

63 21.7 0.0 43.8

88 16.9 0.0 31.6









#35
Dibromofluoromethane
Concen: 49.349 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:113 Resp: 97314
Ion Ratio Lower Upper
113 100
111 99.4 83.3 124.9
192 18.9 13.5 20.3



#36
1,1-Dichloropropene
Concen: 43.735 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 75 Resp: 119608
Ion Ratio Lower Upper
75 100
110 35.9 16.9 50.7
77 31.7 24.7 37.1





#37

Ethyl Acetate

Concen: 49.688 ug/l

RT: 7.553 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 123291

Ion Ratio Lower Upper

43 100

61 14.4 10.7 16.1

70 12.5 8.8 13.2



#38

Carbon Tetrachloride

Concen: 48.253 ug/l

RT: 8.353 min Scan# 1088

Delta R.T. -0.006 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18




Tgt Ion: 117 Resp: 147502


Ion Ratio Lower Upper

117 100

119 95.4 78.7 118.1

121 30.0 24.4 36.6





#39

Methylcyclohexane

Concen: 48.113 ug/l

RT: 9.594 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 133862

Ion Ratio Lower Upper

83 100

55 76.7 62.8 94.2

98 48.9 38.8 58.2



#40

Benzene

Concen: 44.569 ug/l

RT: 8.600 min Scan# 1130

Delta R.T. -0.006 min

Lab File: VN084787.D




Acq: 11 Nov 2024 21:18


Tgt Ion: 78 Resp: 391445

Ion Ratio Lower Upper

78 100

77 22.9 19.1 28.7









#41
Methacrylonitrile
Concen: 51.372 ug/l
RT: 7.771 min Scan# 911
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	68548
Ion	Ratio	Lower	Upper
41	100		
39	55.5	43.3	64.9
67	77.7	56.1	84.1
52	32.4	24.5	36.7



#42
1,2-Dichloroethane
Concen: 46.931 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion:	62	Resp:	133491
Ion	Ratio	Lower	Upper
62	100		
98	9.4	0.0	18.2









#43
Isopropyl Acetate
Concen: 53.170 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 243452
Ion Ratio Lower Upper
43 100
61 24.5 20.4 30.6
87 12.8 10.3 15.5



#44
Trichloroethene
Concen: 45.377 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion:130 Resp: 91914
Ion Ratio Lower Upper
130 100
95 101.8 0.0 203.0





#45

1,2-Dichloropropane

Concen: 46.403 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 95605

Ion Ratio Lower Upper

63 100

65 33.7 24.6 37.0



#46

Dibromomethane

Concen: 46.413 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18




Tgt Ion: 93 Resp: 67476


Ion Ratio Lower Upper

93 100

95 83.0 65.0 97.4

174 96.0 71.8 107.8









#47
Bromodichloromethane
Concen: 47.610 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 145902
Ion Ratio Lower Upper
83 100
85 62.3 51.5 77.3
127 8.6 5.7 8.5#



#48
Methyl methacrylate
Concen: 51.941 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 41 Resp: 100305
Ion Ratio Lower Upper
41 100
69 89.3 67.1 100.7
39 56.9 45.8 68.6





#49

1,4-Dioxane

Concen: 1139.238 ug/l

RT: 9.694 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 38986

Ion Ratio Lower Upper

88 100

43 0.0 25.3 37.9#

58 73.0 56.4 84.6



#50

Toluene-d8

Concen: 48.386 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084787.D




Acq: 11 Nov 2024 21:18


Tgt Ion: 98 Resp: 351476

Ion Ratio Lower Upper

98 100

100 65.3 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 271.309 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 641767
Ion Ratio Lower Upper
43 100
58 40.9 31.6 47.4



#52
Toluene
Concen: 47.993 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 92 Resp: 246524
Ion Ratio Lower Upper
92 100
91 173.0 137.3 205.9









#53
t-1,3-Dichloropropene
Concen: 46.248 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 146707
Ion Ratio Lower Upper
75 100
77 31.9 25.4 38.0



#54
cis-1,3-Dichloropropene
Concen: 45.810 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 75 Resp: 153802
Ion Ratio Lower Upper
75 100
77 30.5 24.2 36.2
39 46.8 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 48.273 ug/l

RT: 11.012 min Scan# 1541

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 93404





Ion Ratio Lower Upper

97 100

83 82.7 71.0 106.4

85 56.2 41.7 62.5

99 60.8 50.1 75.1



#56

Ethyl methacrylate

Concen: 55.974 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18




Tgt Ion: 69 Resp: 156659


Ion Ratio Lower Upper

69 100

41 62.8 53.3 79.9

39 32.9 28.8 43.2





#57

1,3-Dichloropropane

Concen: 48.318 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 160452

Ion Ratio Lower Upper

76 100

78 32.1 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 268.948 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN084787.D




Acq: 11 Nov 2024 21:18


Tgt Ion: 63 Resp: 352227

Ion Ratio Lower Upper

63 100

106 28.9 21.8 32.6









#59
2-Hexanone
Concen: 282.678 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 486706
Ion Ratio Lower Upper
43 100
58 57.0 27.3 81.8



#60
Dibromochloromethane
Concen: 50.662 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 129 Resp: 111721
Ion Ratio Lower Upper
129 100
127 76.1 38.3 114.9









#61
1,2-Dibromoethane
Concen: 48.848 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 96789
Ion Ratio Lower Upper
107 100
109 92.8 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 49.793 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 95 Resp: 135180
Ion Ratio Lower Upper
95 100
174 73.6 0.0 145.2
176 69.9 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 261381
Ion Ratio Lower Upper
117 100
82 55.0 47.2 70.8
119 31.0 25.4 38.0



#64
Tetrachloroethene
Concen: 45.715 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion:164 Resp: 79480
Ion Ratio Lower Upper
164 100
166 125.3 101.0 151.4
129 101.7 80.8 121.2
131 95.2 78.1 117.1









#65
Chlorobenzene
Concen: 45.063 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 263506
Ion Ratio Lower Upper
112 100
114 33.8 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 47.117 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion:131 Resp: 95834
Ion Ratio Lower Upper
131 100
133 94.4 47.9 143.7
119 64.5 32.8 98.3





#67

Ethyl Benzene

Concen: 48.816 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 476489

Ion Ratio Lower Upper

91 100

106 30.8 24.5 36.7



#68

m/p-Xylenes

Concen: 97.406 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN084787.D




Acq: 11 Nov 2024 21:18


Tgt Ion:106 Resp: 359947

Ion Ratio Lower Upper

106 100

91 207.9 167.1 250.7









#69
o-Xylene
Concen: 50.550 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 174747
Ion Ratio Lower Upper
106 100
91 224.0 110.2 330.6



#70
Styrene
Concen: 49.221 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion:104 Resp: 296821
Ion Ratio Lower Upper
104 100
78 52.7 42.7 64.1
103 55.2 43.6 65.4





#71

Bromoform

Concen: 49.894 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:173 Resp: 76363

Ion Ratio Lower Upper

173 100

175 47.4 23.5 70.6

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18




Tgt Ion:152 Resp: 128017


Ion Ratio Lower Upper

152 100

115 63.6 31.3 93.9

150 175.0 0.0 349.8





#73

Isopropylbenzene

Concen: 51.499 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 461999

Ion Ratio Lower Upper

105 100

120 25.7 12.9 38.7



#74

N-ethyl acetate

Concen: 51.881 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Tgt Ion: 43 Resp: 198094




Ion Ratio Lower Upper


43 100

70 45.6 34.4 51.6

55 26.4 19.6 29.4

61 25.9 19.6 29.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 45.679 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN084787.D
 Acq: 11 Nov 2024 21:18




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 83 Resp: 136784
 Ion Ratio Lower Upper
 83 100
 131 10.9 4.9 14.7
 85 65.2 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 69.743 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. 0.006 min
 Lab File: VN084787.D
 Acq: 11 Nov 2024 21:18

Tgt Ion: 75 Resp: 178471
 Ion Ratio Lower Upper
 75 100
 77 130.4 96.9 290.7





#77

Bromobenzene

Concen: 44.282 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.006 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 106796

Ion Ratio Lower Upper

156 100

77 218.0 115.1 345.3

158 98.4 48.9 146.8



#78

n-propylbenzene

Concen: 48.874 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. 0.000 min

Lab File: VN084787.D




Acq: 11 Nov 2024 21:18


Tgt Ion: 91 Resp: 513794

Ion Ratio Lower Upper

91 100

120 22.9 10.8 32.4









#79
2-Chlorotoluene
Concen: 46.294 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 317997
Ion Ratio Lower Upper
91 100
126 32.6 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 50.751 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 105 Resp: 377354
Ion Ratio Lower Upper
105 100
120 47.7 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 50.205 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 54154
Ion Ratio Lower Upper
75 100
53 86.2 78.0 117.0
89 41.1 35.5 53.3



#82
4-Chlorotoluene
Concen: 43.924 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 91 Resp: 317522
Ion Ratio Lower Upper
91 100
126 32.9 15.7 47.1









#83
tert-Butylbenzene
Concen: 54.138 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:119 Resp: 333151
Ion Ratio Lower Upper
119 100
91 68.2 34.1 102.3
134 25.7 12.8 38.4



#84
1,2,4-Trimethylbenzene
Concen: 49.655 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion:105 Resp: 381057
Ion Ratio Lower Upper
105 100
120 44.0 22.3 66.8





#85

sec-Butylbenzene

Concen: 50.669 ug/l

RT: 13.618 min Scan# 1982

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 440437

Ion Ratio Lower Upper

105 100

134 19.7 9.8 29.4



#86

p-Isopropyltoluene

Concen: 52.480 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18




Tgt Ion:119 Resp: 374128


Ion Ratio Lower Upper

119 100

134 25.4 12.6 37.8

91 25.2 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 41.628 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 195927

Ion Ratio Lower Upper

146 100

111 42.5 21.8 65.3

148 63.4 31.9 95.7



#88

1,4-Dichlorobenzene

Concen: 45.560 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18




Tgt Ion:146 Resp: 197206


Ion Ratio Lower Upper

146 100

111 41.7 21.2 63.6

148 62.8 31.9 95.9





#89

n-Butylbenzene

Concen: 46.596 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 310726

Ion Ratio Lower Upper

91 100

92 54.3 26.6 79.7

134 25.6 12.3 36.9



#90

Hexachloroethane

Concen: 43.552 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN084787.D




Acq: 11 Nov 2024 21:18


Tgt Ion: 117 Resp: 69240

Ion Ratio Lower Upper

117 100

201 74.2 34.4 103.2









#91
1,2-Dichlorobenzene
Concen: 44.970 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 203390
Ion Ratio Lower Upper
146 100
111 43.6 22.0 66.0
148 62.9 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 52.063 ug/l
RT: 14.718 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion: 75 Resp: 31583
Ion Ratio Lower Upper
75 100
155 84.5 38.1 114.5
157 104.1 49.6 149.0









#93
1,2,4-Trichlorobenzene
Concen: 47.394 ug/l
RT: 15.394 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:180 Resp: 117324
Ion Ratio Lower Upper
180 100
182 92.7 47.0 141.0
145 33.5 16.4 49.4



#94
Hexachlorobutadiene
Concen: 44.315 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN084787.D
Acq: 11 Nov 2024 21:18

Tgt Ion:225 Resp: 48175
Ion Ratio Lower Upper
225 100
223 65.2 32.4 97.2
227 64.8 32.4 97.0





#95

Naphthalene

Concen: 54.353 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 402741

Ion Ratio Lower Upper

128 100

127 13.0 10.5 15.7

129 11.1 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 47.936 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084787.D

Acq: 11 Nov 2024 21:18

Tgt Ion:180 Resp: 115194

Ion Ratio Lower Upper

180 100

182 92.7 48.2 144.6

145 34.6 17.3 51.9

