

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4774

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN084740.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1108WBS01	Chloromethane	20	14.3	ug/L	72			50	139	
	Vinyl chloride	20	16.2	ug/L	81			58	137	
	Bromomethane	20	17.0	ug/L	85			53	141	
	Chloroethane	20	16.8	ug/L	84			60	138	
	Trichlorofluoromethane	20	17.8	ug/L	89			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.0	ug/L	90			70	136	
	1,1-Dichloroethene	20	16.7	ug/L	84			71	131	
	Acetone	100	110	ug/L	110			39	160	
	Carbon disulfide	20	15.1	ug/L	76			64	133	
	Methyl tert-butyl Ether	20	18.9	ug/L	95			71	124	
	Methylene Chloride	20	17.8	ug/L	89			74	124	
	trans-1,2-Dichloroethene	20	16.8	ug/L	84			75	124	
	1,1-Dichloroethane	20	17.9	ug/L	90			77	125	
	2-Butanone	100	100	ug/L	100			56	143	
	Carbon Tetrachloride	20	19.0	ug/L	95			72	136	
	cis-1,2-Dichloroethene	20	18.3	ug/L	92			78	123	
	Chloroform	20	18.8	ug/L	94			79	124	
	1,1,1-Trichloroethane	20	18.7	ug/L	94			74	131	
	Methylcyclohexane	20	18.5	ug/L	93			72	132	
	Benzene	20	17.7	ug/L	89			79	120	
	1,2-Dichloroethane	20	19.1	ug/L	96			73	128	
	Trichloroethene	20	18.3	ug/L	92			79	123	
	1,2-Dichloropropane	20	18.6	ug/L	93			78	122	
	Bromodichloromethane	20	18.3	ug/L	92			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	18.9	ug/L	95			80	121	
	t-1,3-Dichloropropene	20	18.2	ug/L	91			73	127	
	cis-1,3-Dichloropropene	20	18.3	ug/L	92			75	124	
	1,1,2-Trichloroethane	20	19.6	ug/L	98			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	20.0	ug/L	100			74	126	
	Tetrachloroethene	20	18.3	ug/L	92			74	129	
	Chlorobenzene	20	17.6	ug/L	88			82	118	
	Ethyl Benzene	20	18.2	ug/L	91			79	121	
	m/p-Xylenes	40	36.9	ug/L	92			80	121	
	o-Xylene	20	18.8	ug/L	94			78	122	
	Styrene	20	18.6	ug/L	93			78	123	
	Bromoform	20	19.8	ug/L	99			66	130	
	Isopropylbenzene	20	17.4	ug/L	87			72	131	
	1,1,2,2-Tetrachloroethane	20	18.1	ug/L	91			71	121	
	1,3-Dichlorobenzene	20	16.0	ug/L	80			80	119	
	1,4-Dichlorobenzene	20	16.5	ug/L	83			79	118	
	1,2-Dichlorobenzene	20	16.8	ug/L	84			80	119	