

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4774 SAS No.: P4774 SDG NO.: P4774
Lab File ID: VN084737.D Date Analyzed: 11/08/2024
Instrument ID: MSVOA_N Time Analyzed: 09:39
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	135765	13.788				
UPPER LIMIT	271530	14.288				
LOWER LIMIT	67882.5	13.288				
EPA SAMPLE NO.						
BP-VPB-190-TB-20241105	111909	13.79				
VPB190-HYD-20241106	131449	13.79				
BP-VPB-190-EB-20241106	112988	13.79				
BP-VPB-190-DUP-20241105	123463	13.79				
BP-VPB-190-GW-618-620	121391	13.79				
BP-VPB-190-GW-638-640	129789	13.79				
BP-VPB-190-GW-658-660	129510	13.79				
VN1108WBL01	113797	13.79				
VN1108WBS01	135553	13.79				
VN1108WBSD01	120765	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.