

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4774 SAS No.: P4774 SDG NO.: P4774
Lab File ID: VN084569.D BFB Injection Date: 10/30/2024
Instrument ID: MSVOA_N BFB Injection Time: 10:42
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.5
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.2
173	Less than 2.0% of mass 174	1.2 (1.6) 1
174	50.0 - 100.0% of mass 95	73.5
175	5.0 - 9.0% of mass 174	5.7 (7.7) 1
176	95.0 - 101.0% of mass 174	70.1 (95.4) 1
177	5.0 - 9.0% of mass 176	4.8 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN084570.D	10/30/2024	11:46
VSTDICCC050	VSTDICCC050	VN084571.D	10/30/2024	12:09
VSTDICC020	VSTDICC020	VN084572.D	10/30/2024	12:33
VSTDICC010	VSTDICC010	VN084573.D	10/30/2024	12:57
VSTDICC005	VSTDICC005	VN084574.D	10/30/2024	13:21
VSTDICC001	VSTDICC001	VN084575.D	10/30/2024	13:45