

Surrogate Summary

SDG No.: **P4779**

Client: **CHA Companies, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP068360.D	PIBLK-PP068360.D	Tetrachloro-m-xylene	1	20	21.0	105		60	140
		Decachlorobiphenyl	1	20	23.4	117		60	140
		Tetrachloro-m-xylene	2	20	21.6	108		60	140
		Decachlorobiphenyl	2	20	23.6	118		60	140
I.BLK-PP068399.D	PIBLK-PP068399.D	Tetrachloro-m-xylene	1	20	22.1	110		60	140
		Decachlorobiphenyl	1	20	23.3	117		60	140
		Tetrachloro-m-xylene	2	20	21.7	109		60	140
		Decachlorobiphenyl	2	20	23.9	120		60	140
PB164860BL	PB164860BL	Tetrachloro-m-xylene	1	20	19.7	98		32	144
		Decachlorobiphenyl	1	20	18.4	92		32	175
		Tetrachloro-m-xylene	2	20	20.4	102		32	144
		Decachlorobiphenyl	2	20	18.3	91		32	175
PB164860BS	PB164860BS	Tetrachloro-m-xylene	1	20	20.7	103		32	144
		Decachlorobiphenyl	1	20	19.0	95		32	175
		Tetrachloro-m-xylene	2	20	21.8	109		32	144
		Decachlorobiphenyl	2	20	20.2	101		32	175
P4779-01	SOIL-01	Tetrachloro-m-xylene	1	20	20.5	103		32	144
		Decachlorobiphenyl	1	20	23.0	115		32	175
		Tetrachloro-m-xylene	2	20	19.7	99		32	144
		Decachlorobiphenyl	2	20	22.7	114		32	175
I.BLK-PP068414.D	PIBLK-PP068414.D	Tetrachloro-m-xylene	1	20	21.6	108		60	140
		Decachlorobiphenyl	1	20	22.7	113		60	140
		Tetrachloro-m-xylene	2	20	21.9	109		60	140
		Decachlorobiphenyl	2	20	23.3	117		60	140
P4789-01MS	BP-F21MS	Tetrachloro-m-xylene	1	20	24.5	123		32	144
		Decachlorobiphenyl	1	20	22.5	113		32	175
		Tetrachloro-m-xylene	2	20	23.2	116		32	144
		Decachlorobiphenyl	2	20	23.1	115		32	175
P4789-01MSD	BP-F21MSD	Tetrachloro-m-xylene	1	20	24.7	124		32	144
		Decachlorobiphenyl	1	20	22.9	114		32	175
		Tetrachloro-m-xylene	2	20	25.2	126		32	144
		Decachlorobiphenyl	2	20	22.9	114		32	175
I.BLK-PP068429.D	PIBLK-PP068429.D	Tetrachloro-m-xylene	1	20	21.8	109		60	140
		Decachlorobiphenyl	1	20	22.2	111		60	140
		Tetrachloro-m-xylene	2	20	22.3	111		60	140
		Decachlorobiphenyl	2	20	23.0	115		60	140