

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CLOU03
Lab Code: CHEM Case No.: P4779 SAS No.: P4779 SDG No.: P4779
Instrument ID: BNA_F Calibration Date(s): 11/13/2024 11/13/2024
Calibration Time(s): 09:01 12:48

LAB FILE ID:		RRF2.5 = BF140332.D			RRF005 = BF140333.D		RRF010 = BF140334.D	
		RRF020 = BF140335.D			RRF050 = BF140337.D		RRF060 = BF140338.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF050	RRF060	RRF	% RSD
2-Fluorophenol		1.329	1.257	1.223	1.090	1.118	1.171	8.8
Benzaldehyde		1.101	1.087	1.017	0.828	0.783	0.951	14.3
Phenol-d6		1.773	1.691	1.643	1.483	1.507	1.587	8.1
Phenol		1.799	1.793	1.743	1.582	1.597	1.674	7.3
bis(2-Chloroethyl) ether		1.359	1.294	1.306	1.227	1.249	1.268	4.6
2-Chlorophenol		1.397	1.349	1.332	1.176	1.184	1.258	8.5
2-Methylphenol		1.092	1.062	1.072	0.998	1.016	1.035	5.1
2,2-oxybis(1-Chloropropane)		1.906	1.826	1.815	1.666	1.684	1.752	7.0
Acetophenone		0.532	0.508	0.481	0.434	0.441	0.465	9.4
3+4-Methylphenols		1.436	1.427	1.359	1.200	1.198	1.295	10.2
n-Nitroso-di-n-propylamine	1.032	1.029	1.003	0.986	0.883	0.905	0.959	7.3
Nitrobenzene-d5		0.411	0.401	0.399	0.368	0.379	0.384	5.3
Hexachloroethane		0.562	0.543	0.550	0.500	0.506	0.520	6.3
Nitrobenzene		0.423	0.421	0.406	0.390	0.394	0.400	4.6
Isophorone		0.736	0.702	0.692	0.654	0.671	0.680	4.9
2-Nitrophenol		0.175	0.179	0.181	0.175	0.181	0.177	2.0
2,4-Dimethylphenol		0.238	0.236	0.232	0.221	0.227	0.227	3.9
bis(2-Chloroethoxy) methane		0.461	0.442	0.430	0.398	0.404	0.417	6.8
2,4-Dichlorophenol		0.307	0.296	0.287	0.272	0.272	0.281	6.2
Naphthalene		1.160	1.105	1.064	0.950	0.955	1.015	9.6
4-Chloroaniline		0.390	0.386	0.370	0.333	0.340	0.353	8.4
Hexachlorobutadiene		0.220	0.214	0.210	0.188	0.190	0.199	7.7
Caprolactam		0.098	0.093	0.094	0.091	0.095	0.093	3.3
4-Chloro-3-methylphenol		0.331	0.327	0.320	0.299	0.304	0.311	5.1
2-Methylnaphthalene		0.751	0.706	0.683	0.610	0.607	0.651	10.0
Hexachlorocyclopentadiene			0.111	0.143	0.156	0.159	0.142	12.0
2,4,6-Trichlorophenol		0.380	0.374	0.374	0.360	0.352	0.363	3.9
2-Fluorobiphenyl		1.524	1.396	1.351	1.125	1.110	1.244	14.5
2,4,5-Trichlorophenol		0.421	0.417	0.412	0.387	0.390	0.397	5.4
1,1-Biphenyl		1.690	1.571	1.565	1.368	1.343	1.455	10.8
2-Chloronaphthalene		1.279	1.182	1.149	1.064	1.049	1.108	9.2
2-Nitroaniline		0.378	0.365	0.384	0.365	0.366	0.368	2.9
Dimethylphthalate		1.464	1.358	1.371	1.246	1.244	1.304	7.5
Acenaphthylene		1.940	1.802	1.786	1.582	1.552	1.677	10.3

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CLOU03
Lab Code: CHEM Case No.: P4779 SAS No.: P4779 SDG No.: P4779
Instrument ID: BNA_F Calibration Date(s): 11/13/2024 11/13/2024
Calibration Time(s): 09:01 12:48

LAB FILE ID:		RRF2.5 = BF140332.D			RRF005 = BF140333.D		RRF010 = BF140334.D	
		RRF020 = BF140335.D			RRF050 = BF140337.D		RRF060 = BF140338.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF050	RRF060	RRF	% RSD
2,6-Dinitrotoluene		0.317	0.303	0.313	0.291	0.289	0.297	5.6
3-Nitroaniline		0.333	0.327	0.335	0.307	0.298	0.312	6.9
Acenaphthene		1.259	1.204	1.192	1.091	1.079	1.133	7.8
2,4-Dinitrophenol			0.109	0.148	0.178	0.170	0.153	16.3
4-Nitrophenol		0.199	0.217	0.239	0.240	0.237	0.228	6.7
Dibenzofuran		1.904	1.736	1.715	1.502	1.466	1.603	11.9
2,4-Dinitrotoluene		0.414	0.400	0.417	0.387	0.382	0.391	5.8
Diethylphthalate		1.525	1.417	1.398	1.272	1.246	1.333	8.8
4-Chlorophenyl-phenylether		0.739	0.688	0.669	0.583	0.567	0.625	12.0
Fluorene		1.525	1.403	1.365	1.168	1.137	1.267	13.1
4-Nitroaniline		0.335	0.325	0.328	0.316	0.313	0.319	3.9
4,6-Dinitro-2-methylphenol		0.080	0.098	0.113	0.117	0.120	0.108	13.0
n-Nitrosodiphenylamine		0.651	0.630	0.600	0.541	0.541	0.578	8.6
2,4,6-Tribromophenol		0.220	0.211	0.207	0.191	0.189	0.199	6.9
4-Bromophenyl-phenylether		0.227	0.210	0.209	0.191	0.188	0.200	8.1
Hexachlorobenzene		0.251	0.242	0.232	0.209	0.215	0.225	7.5
Atrazine		0.191	0.183	0.135	0.155	0.205	0.175	17.0
Pentachlorophenol		0.097	0.108	0.124	0.131	0.130	0.121	10.9
Phenanthrene		1.096	1.053	0.986	0.868	0.851	0.940	11.3
Anthracene		1.071	1.018	0.966	0.860	0.839	0.921	10.8
Carbazole		1.053	1.009	0.963	0.846	0.834	0.909	11.0
Di-n-butylphthalate		1.214	1.177	1.150	1.036	1.023	1.091	8.4
Fluoranthene		1.256	1.201	1.141	0.962	0.933	1.054	13.9
Pyrene		1.748	1.706	1.643	1.728	1.800	1.711	3.1
Terphenyl-d14		1.226	1.185	1.108	1.139	1.190	1.152	4.3
Butylbenzylphthalate		0.640	0.638	0.654	0.670	0.681	0.657	3.6
3,3-Dichlorobenzidine		0.438	0.431	0.433	0.406	0.416	0.418	3.8
Benzo (a) anthracene		1.451	1.418	1.351	1.229	1.294	1.314	7.3
Chrysene		1.394	1.241	1.235	1.137	1.123	1.199	8.4
Bis (2-ethylhexyl) phthalate		0.886	0.890	0.897	0.858	0.869	0.877	4.1
Di-n-octyl phthalate		1.231	1.217	1.270	1.187	1.241	1.235	2.9
Benzo (b) fluoranthene		1.405	1.352	1.471	1.235	1.233	1.308	7.8
Benzo (k) fluoranthene		1.286	1.241	1.120	0.952	0.935	1.069	14.5
Benzo (a) pyrene		1.100	1.053	1.054	0.970	0.980	1.016	5.4
Indeno (1,2,3-cd) pyrene		1.116	1.150	1.138	1.316	1.354	1.235	8.0

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CLOU03
Lab Code: CHEM Case No.: P4779 SAS No.: P4779 SDG No.: P4779
Instrument ID: BNA_F Calibration Date(s): 11/13/2024 11/13/2024
Calibration Time(s): 09:01 12:48

LAB FILE ID:		RRF2.5 = BF140332.D		RRF005 = BF140333.D		RRF010 = BF140334.D		
		RRF020 = BF140335.D		RRF050 = BF140337.D		RRF060 = BF140338.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF050	RRF060	RRF	% RSD
Dibenzo (a,h) anthracene		0.941	0.952	0.944	1.074	1.121	1.021	7.3
Benzo (g,h,i) perylene		0.941	0.963	0.954	1.122	1.146	1.044	8.5
1,2,4,5-Tetrachlorobenzene		0.646	0.585	0.583	0.524	0.517	0.553	9.8
1,4-Dioxane		0.582	0.546	0.530	0.495	0.506	0.522	6.5
2,3,4,6-Tetrachlorophenol		0.322	0.328	0.326	0.308	0.302	0.313	4.6