

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110624

Review By	Jagrut	Review On	11/7/2024 11:03:46 AM
Supervise By	mohammad	Supervise On	11/8/2024 3:53:54 AM
SubDirectory	BE110624	HP Acquire Method	BNA_E
		HP Processing Method	be110624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12324,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101491.D	6 Nov 2024 11:40		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BE101492.D	6 Nov 2024 13:15	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BE101493.D	6 Nov 2024 13:51		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BE101494.D	6 Nov 2024 14:27	Compound #32,54 removed from 5ppm	RC/JU	Ok,M
5	SSTDICC010	SSTDICC010	BE101495.D	6 Nov 2024 15:03		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BE101496.D	6 Nov 2024 15:39	Compound #32 Kept on LR	RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BE101497.D	6 Nov 2024 16:14	The Calibration is Good For 8270 DOD Except com#77 and will NOT be used for 625.1 Method	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BE101498.D	6 Nov 2024 16:50	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
9	SSTDICC060	SSTDICC060	BE101499.D	6 Nov 2024 17:26		RC/JU	Ok,M
10	SSTDICC080	SSTDICC080	BE101500.D	6 Nov 2024 18:02		RC/JU	Ok,M
11	SSTDICV040	ICVBE110624	BE101501.D	6 Nov 2024 18:41		RC/JU	Ok,M
12	PB164561BL	PB164561BL	BE101502.D	6 Nov 2024 19:16		RC/JU	Ok
13	DFTPP	DFTPP	BE101503.D	6 Nov 2024 19:52		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BE101504.D	6 Nov 2024 20:28		RC/JU	Ok,M
15	PB164561TB	PB164561TB	BE101505.D	6 Nov 2024 21:04		RC/JU	Ok
16	PB164561BS	PB164561BS	BE101506.D	6 Nov 2024 21:40		RC/JU	Ok,M

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110624

Review By	Jagrut	Review On	11/7/2024 11:03:46 AM		
Supervise By	mohammad	Supervise On	11/8/2024 3:53:54 AM		
SubDirectory	BE110624	HP Acquire Method	BNA_E	HP Processing Method	be110624
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12324,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4660-11	WC-CONCRETE-01-C	BE101507.D	6 Nov 2024 22:15		RC/JU	Ok
18	P4667-04	BP-F-6	BE101508.D	6 Nov 2024 22:51		RC/JU	Ok
19	P4667-08	BP-F-5	BE101509.D	6 Nov 2024 23:27	Internal Standard and Surrogate fail	RC/JU	ReRun
20	P4667-12	TP-10	BE101510.D	7 Nov 2024 00:03	Internal Standard Fail	RC/JU	ReRun
21	P4667-16	BP-F-7	BE101511.D	7 Nov 2024 00:39		RC/JU	Ok
22	P4659-04	MH-2	BE101512.D	7 Nov 2024 1:14		RC/JU	Ok
23	P4679-04	MH-1	BE101513.D	7 Nov 2024 1:50		RC/JU	Ok
24	P4680-04	BP-F26	BE101514.D	7 Nov 2024 2:26	Internal Standrad Fail	RC/JU	ReRun
25	P4680-08	BP-F25	BE101515.D	7 Nov 2024 3:02	Internal Standrad Fail	RC/JU	ReRun
26	P4660-07	WC-WOOD-01-C	BE101516.D	7 Nov 2024 3:38		RC/JU	Ok,M
27	P4711-05	CF-613-COMP-16	BE101517.D	7 Nov 2024 4:13	Internal Standrad Fail	RC/JU	ReRun
28	P4711-10	CF-613-COMP-17	BE101518.D	7 Nov 2024 4:49	Internal Standrad Fail	RC/JU	ReRun
29	P4711-01	CF-613-COMP-16	BE101519.D	7 Nov 2024 5:25		RC/JU	Ok,M
30	P4711-06	CF-613-COMP-17	BE101520.D	7 Nov 2024 6:01	Need further 5X Dilution	RC/JU	Dilution
31	P4708-01	OR-02-110424	BE101521.D	7 Nov 2024 6:37		RC/JU	Ok,M
32	P4706-01	TR-04-110424	BE101522.D	7 Nov 2024 7:13		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE111124

Review By	yogesh	Review On	11/12/2024 1:19:10 AM
Supervise By	mohammad	Supervise On	11/12/2024 1:26:15 AM
SubDirectory	BE111124	HP Acquire Method	BNA_E
		HP Processing Method	be110624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12326,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101563.D	11 Nov 2024 9:43		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BE101564.D	11 Nov 2024 10:19		RC/JU	Ok,M
3	PB164845BL	PB164845BL	BE101565.D	11 Nov 2024 10:55		RC/JU	Ok
4	PB164845BS	PB164845BS	BE101566.D	11 Nov 2024 11:31		RC/JU	Ok,M
5	P4756-04	BP-B4	BE101567.D	11 Nov 2024 12:09		RC/JU	Ok
6	P4756-04MS	BP-B4MS	BE101568.D	11 Nov 2024 12:45		RC/JU	Ok,M
7	P4756-04MSD	BP-B4MSD	BE101569.D	11 Nov 2024 13:21		RC/JU	Ok,M
8	P4793-02	M00-24-00345	BE101570.D	11 Nov 2024 13:57		RC/JU	Ok
9	P4795-02	LAW-23-00189	BE101571.D	11 Nov 2024 14:33		RC/JU	Ok
10	P4796-07	TP-4	BE101572.D	11 Nov 2024 15:08	Internal Standard Fail	RC/JU	ReRun
11	P4796-03	TP-2	BE101573.D	11 Nov 2024 15:44		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111324

Review By	yogesh	Review On	11/14/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:17 AM
SubDirectory	BF111324	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140331.D	13 Nov 2024 08:35		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140332.D	13 Nov 2024 09:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140333.D	13 Nov 2024 09:27	Compound#41,54,77 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF140334.D	13 Nov 2024 09:53		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF140335.D	13 Nov 2024 10:29		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF140336.D	13 Nov 2024 10:55	Injection Error	RC/JU	Not Ok
7	SSTDICC050	SSTDICC050	BF140337.D	13 Nov 2024 11:21		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF140338.D	13 Nov 2024 11:47		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF140339.D	13 Nov 2024 12:13	Compound#09 removed from 80 ppm	RC/JU	Ok,M
10	SSTDICCC040	SSTDICCC040	BF140340.D	13 Nov 2024 12:48		RC/JU	Ok
11	SSTDICV040	ICVBF111324	BF140341.D	13 Nov 2024 13:19		RC/JU	Ok
12	PB164401BL	PB164401BL	BF140342.D	13 Nov 2024 13:45		RC/JU	Ok
13	P4495-11	PT-BNA-SOIL	BF140343.D	13 Nov 2024 15:25	PT Sample, Need 5X Dilution	RC/JU	Dilution
14	P4495-11DL	PT-BNA-SOILDL	BF140344.D	13 Nov 2024 16:02		RC/JU	Ok
15	P3845-03	PT-BN-WP	BF140345.D	13 Nov 2024 16:40	PT Sample Analyzed for confirmation - Need 5X Dilution.	RC/JU	Dilution
16	P3845-03DL	PT-BN-WPDL	BF140346.D	13 Nov 2024 17:06		RC/JU	Ok
17	P3845-06	PT-ACIDS-WP	BF140347.D	13 Nov 2024 17:32	PT Sample Analyzed for confirmation - Need 5X Dilution.	RC/JU	Dilution

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111324

Review By	yogesh	Review On	11/14/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:17 AM
SubDirectory	BF111324	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	P3845-06DL	PT-ACIDS-WPDL	BF140348.D	13 Nov 2024 17:58	Dilution not matching (Low results)	RC/JU	Not Ok
19	SP	SP	BF140349.D	13 Nov 2024 18:39	8270-625.1 SPIKE SOLUTION	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111424

Review By	yogesh	Review On	11/15/2024 5:15:55 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:37 AM
SubDirectory	BF111424	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140350.D	14 Nov 2024 08:47		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140351.D	14 Nov 2024 09:42		RC/JU	Ok
3	PB164908BL	PB164908BL	BF140352.D	14 Nov 2024 10:08		RC/JU	Ok
4	PB164911BS	PB164911BS	BF140353.D	14 Nov 2024 10:34		RC/JU	Ok,M
5	PB164911BSD	PB164911BSD	BF140354.D	14 Nov 2024 11:01		RC/JU	Ok,M
6	PB164911BL	PB164911BL	BF140355.D	14 Nov 2024 11:27		RC/JU	Ok
7	P3845-06DL	PT-ACIDS-WPDL	BF140356.D	14 Nov 2024 11:54		RC/JU	Ok
8	P4762-02	EFF-WW	BF140357.D	14 Nov 2024 13:03		RC/JU	Ok
9	P4792-03	OILY-STONE-DRUM	BF140358.D	14 Nov 2024 13:29		RC/JU	Ok
10	P4837-03	3TRK-STONE	BF140359.D	14 Nov 2024 13:55		RC/JU	Ok
11	P4834-03	SLP-1-STONE	BF140360.D	14 Nov 2024 14:21		RC/JU	Ok
12	P4779-01	SOIL-01	BF140361.D	14 Nov 2024 14:47		RC/JU	Ok
13	P4838-03	T1-STONE	BF140362.D	14 Nov 2024 15:13		RC/JU	Ok
14	P4718-04	WB-307-SW	BF140363.D	14 Nov 2024 15:39		RC/JU	Ok
15	P4812-01	RBR200035	BF140364.D	14 Nov 2024 16:04		RC/JU	Ok
16	DFTPP	DFTPP	BF140365.D	14 Nov 2024 16:35		RC/JU	Ok
17	SSTDCCC040	SSTDCCC040	BF140366.D	14 Nov 2024 17:02		RC/JU	Ok
18	PB164846BL	PB164846BL	BF140367.D	14 Nov 2024 17:28		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111424

Review By	yogesh	Review On	11/15/2024 5:15:55 AM				
Supervise By	mohammad	Supervise On	11/15/2024 5:49:37 AM				
SubDirectory	BF111424	HP Acquire Method	BNA_F	HP Processing Method	Bf111324		
STD. NAME		STD REF.#					
Tune/Reschk		SP6573					
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621					
CCC		SP6624					
Internal Standard/PEM		S12327,10ul/1000ul sample					
ICV/I.BLK		SP6559					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

19	P4722-10	WC-2(0-6)	BF140368.D	14 Nov 2024 17:59		RC/JU	Ok
20	P4722-15	WC-3(0-6)	BF140369.D	14 Nov 2024 18:25		RC/JU	Ok
21	P4791-02	HINCHMAN-OILY-WAT	BF140370.D	14 Nov 2024 18:51	Internal Standard Fail	RC/JU	Ok
22	P4792-04	MANHOLE-WASTE-DR	BF140371.D	14 Nov 2024 19:17		RC/JU	Ok
23	P4788-01	BP-G3	BF140372.D	14 Nov 2024 19:43		RC/JU	Ok
24	P4788-01MS	BP-G3MS	BF140373.D	14 Nov 2024 20:09		RC/JU	Ok
25	P4788-01MSD	BP-G3MSD	BF140374.D	14 Nov 2024 20:35		RC/JU	Ok
26	P4796-05RE	TP-3RE	BF140375.D	14 Nov 2024 21:01	Fax is already given	RC/JU	Confirms
27	P4796-01RE	TP-1RE	BF140376.D	14 Nov 2024 21:28	Fax is already given	RC/JU	Confirms
28	P4722-08DL	WC-2(0-6)DL	BF140377.D	14 Nov 2024 21:54		RC/JU	Ok,M
29	P4768-06DL	B-6-2DL	BF140378.D	14 Nov 2024 22:20		RC/JU	Ok,M
30	P4739-09DL	BP-B2DL	BF140379.D	14 Nov 2024 22:47		RC/JU	Ok
31	P4795-01RE	LAW-23-00189RE	BF140380.D	14 Nov 2024 23:13	Fax is already given	RC/JU	Confirms
32	P4811-01RE	RB24008RE	BF140381.D	14 Nov 2024 23:39	Internal Standard Fail	RC/JU	Confirms
33	P4816-01RE	331RE	BF140382.D	15 Nov 2024 00:05	Fax is already given	RC/JU	Confirms
34	P4807-05	BF-F14	BF140383.D	15 Nov 2024 00:32		RC/JU	Ok
35	P4807-05MS	BF-F14MS	BF140384.D	15 Nov 2024 00:58		RC/JU	Ok
36	P4807-05MSD	BF-F14MSD	BF140385.D	15 Nov 2024 01:24		RC/JU	Ok
37	P4807-01	TP-7	BF140386.D	15 Nov 2024 01:49		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111424

Review By	yogesh	Review On	11/15/2024 5:15:55 AM		
Supervise By	mohammad	Supervise On	11/15/2024 5:49:37 AM		
SubDirectory	BF111424	HP Acquire Method	BNA_F	HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

38	P4798-01	MH-6	BF140387.D	15 Nov 2024 02:16	Need 2X Dilution	RC/JU	Dilution
39	P4789-01	BP-F21	BF140388.D	15 Nov 2024 02:42		RC/JU	Ok
40	P4793-01	M00-24-00345	BF140389.D	15 Nov 2024 03:08		RC/JU	Confirms

M : Manual Integration