

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4779 SAS No.: P4779 SDG NO.: P4779

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/11/2024

Lab File ID: BE101564.D Time Analyzed: 10:19

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	67067	7.559	291057	10.33	192797	14.17
UPPER LIMIT	134134	8.059	582114	10.826	385594	14.669
LOWER LIMIT	33533.5	7.059	145529	9.826	96398.5	13.669
EPA SAMPLE NO.						
01 PB164845BL	54855	7.56	227680	10.33	136558	14.17
02 PB164845BS	62806	7.56	266549	10.32	161238	14.17

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.