

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CLOU03

Lab Code: CHEM

SAS No.: P4779 SDG NO.: P4779

Lab File ID: BE101491.D

DFTPP Injection Date: 11/06/2024

Instrument ID: BNA_E

DFTPP Injection Time: 11:40

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	14.8
68	Less than 2.0% of mass 69	0.2 (1.3) 1
69	Mass 69 relative abundance	16
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	23.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	3.6
275	10.0 - 60.0% of mass 198	18.3
365	Greater than 1% of mass 198	2.9
441	Present, but less than mass 443	16.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BE101493.D	11/06/2024	13:51
SSTDICC005	SSTDICC005	BE101494.D	11/06/2024	14:27
SSTDICC010	SSTDICC010	BE101495.D	11/06/2024	15:03
SSTDICC020	SSTDICC020	BE101496.D	11/06/2024	15:39
SSTDICCC040	SSTDICCC040	BE101497.D	11/06/2024	16:14
SSTDICC050	SSTDICC050	BE101498.D	11/06/2024	16:50
SSTDICC060	SSTDICC060	BE101499.D	11/06/2024	17:26
SSTDICC080	SSTDICC080	BE101500.D	11/06/2024	18:02