

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4780 SAS No.: P4780 SDG NO.: P4780

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 10/29/2024 10/29/2024

Continuing Calib Time: 15:07 Initial Calibration Time(s): 04:56 14:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.86	4.87	4.80	4.94	0.01
Aroclor-1242-2	(2)	4.88	4.89	4.82	4.96	0.01
Aroclor-1242-3	(3)	4.95	4.95	4.88	5.02	0.00
Aroclor-1242-4	(4)	5.05	5.06	4.99	5.13	0.01
Aroclor-1242-5	(5)	5.84	5.84	5.77	5.91	0.00
TCX		3.64	3.64	3.59	3.69	0.00
DCB		9.53	9.54	9.44	9.64	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4780 SAS No.: P4780 SDG NO.: P4780

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 10/29/2024 10/29/2024

Continuing Calib Time: 15:07 Initial Calibration Time(s): 04:56 14:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.09	4.10	4.03	4.17	0.01
Aroclor-1242-2	(2)	4.11	4.12	4.05	4.19	0.01
Aroclor-1242-3	(3)	4.29	4.30	4.23	4.37	0.01
Aroclor-1242-4	(4)	4.39	4.39	4.32	4.46	0.00
Aroclor-1242-5	(5)	4.94	4.94	4.87	5.01	0.01
TCX		2.97	2.97	2.92	3.02	0.00
DCB		8.26	8.26	8.16	8.36	0.00