

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4780 SAS No.: P4780 SDG NO.: P4780

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 10/29/2024 10/29/2024

Continuing Calib Time: 20:01 Initial Calibration Time(s): 04:56 14:43

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1 | (1) | 4.86       | 4.87      | 4.80                   | 4.94 | 0.01       |
| Aroclor-1016-2 | (2) | 4.88       | 4.89      | 4.82                   | 4.96 | 0.01       |
| Aroclor-1016-3 | (3) | 4.95       | 4.95      | 4.88                   | 5.02 | 0.00       |
| Aroclor-1016-4 | (4) | 5.05       | 5.06      | 4.99                   | 5.13 | 0.01       |
| Aroclor-1016-5 | (5) | 5.36       | 5.37      | 5.30                   | 5.44 | 0.01       |
| Aroclor-1260-1 | (1) | 6.58       | 6.58      | 6.51                   | 6.65 | 0.00       |
| Aroclor-1260-2 | (2) | 6.86       | 6.86      | 6.79                   | 6.93 | 0.00       |
| Aroclor-1260-3 | (3) | 7.25       | 7.25      | 7.18                   | 7.32 | 0.00       |
| Aroclor-1260-4 | (4) | 7.49       | 7.50      | 7.43                   | 7.57 | 0.01       |
| Aroclor-1260-5 | (5) | 7.83       | 7.83      | 7.76                   | 7.90 | 0.00       |
| TCX            |     | 3.64       | 3.65      | 3.60                   | 3.70 | 0.01       |
| DCB            |     | 9.53       | 9.53      | 9.43                   | 9.63 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4780 SAS No.: P4780 SDG NO.: P4780

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 10/29/2024 10/29/2024

Continuing Calib Time: 20:01 Initial Calibration Time(s): 04:56 14:43

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------|-----|------------|-----------|-----------|------|------------|
|                |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1 | (1) | 4.09       | 4.10      | 4.03      | 4.17 | 0.01       |
| Aroclor-1016-2 | (2) | 4.11       | 4.12      | 4.05      | 4.19 | 0.01       |
| Aroclor-1016-3 | (3) | 4.29       | 4.30      | 4.23      | 4.37 | 0.01       |
| Aroclor-1016-4 | (4) | 4.34       | 4.35      | 4.28      | 4.42 | 0.01       |
| Aroclor-1016-5 | (5) | 4.56       | 4.57      | 4.50      | 4.64 | 0.01       |
| Aroclor-1260-1 | (1) | 5.66       | 5.67      | 5.60      | 5.74 | 0.01       |
| Aroclor-1260-2 | (2) | 5.87       | 5.87      | 5.80      | 5.94 | 0.00       |
| Aroclor-1260-3 | (3) | 6.03       | 6.03      | 5.96      | 6.10 | 0.00       |
| Aroclor-1260-4 | (4) | 6.53       | 6.54      | 6.47      | 6.61 | 0.01       |
| Aroclor-1260-5 | (5) | 6.80       | 6.80      | 6.73      | 6.87 | 0.00       |
| TCX            |     | 2.97       | 2.98      | 2.93      | 3.03 | 0.01       |
| DCB            |     | 8.26       | 8.26      | 8.16      | 8.36 | 0.00       |