

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4780 SAS No.: P4780 SDG NO.: P4780

EPA Sample No.: SSTD020533 Date Analyzed: 11/13/2024

Lab File ID: BG063358.D Time Analyzed: 17:26

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	92534	7.832	390044	10.62	329053	14.47
UPPER LIMIT	185068	8.332	780088	11.123	658106	14.971
LOWER LIMIT	46267	7.332	195022	10.123	164527	13.971
EPA SAMPLE NO.						
01 SVOC-GPC-BLANK	89778	7.83	366144	10.62	326836	14.47
02 SVOC-GPC2-BLANK	79040	7.83	337242	10.62	305323	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.