

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51863 MA No. : SDG No.: MC0DA8
 Initial Calibration Verification Source : MP83142
 Continuing Calibration Verification Source : MP83145
 Run Batch: LB133473 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Lead	1000	990	99	0	25000	25000	98	0	25000	98	0