

## Pratik Patel

---

**From:** Danganan, Eloise B <Eloise.Danganan@gdit.com>  
**Sent:** Wednesday, November 20, 2024 10:39 AM  
**To:** Sohil Jodhani; Mohammad Ahmed; Nimisha Pandya  
**Cc:** R5 - R5RSCC@epa.gov (R5RSCC@epa.gov)  
**Subject:** REGION 5 | CASE 51834 | LAB ACE | SDG E2117 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL  
**Attachments:** P4785-10(E27M1).pdf; P4785-16(E27N7).pdf; P4785-17(E27N8).pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning ACE,

Please see the Region's response regarding the laboratory's reported issue for Case 51834/SDG E2117 and let me know if you have any questions. Also, please be sure to include detailed information about this issue in the SDG Narrative.

**ISSUE:** "Lab has received water samples for VOA analysis. Lab has performed the analysis for the samples E27M1, E27N7 & E27N8 in a continuous analytical sequence. Samples are found positive with high concentration of multiple analysis and required dilution analysis to bring target analytes within calibration range. In this case, instrument blank was analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted VOA analysis without instrument blank in between and further dilution in final electronic data.

Please see attachment for your reference."

**REGION:** "Per the client, the lab's proposal is acceptable."

\*\*\*\*\* Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

ELOISE DANGANAN  
Environmental Analyst  
GDIT Federal Civilian Division  
Under contract to the EPA  
T 703.254.0067  
15036 Conference Center Drive  
Chantilly, VA 20151  
[www.gdit.com](http://www.gdit.com)

**GENERAL DYNAMICS**  
Information Technology

LEAVE ALERT: 11/22-11/25, 11/28

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE:

---

**From:** R5RSCC <R5RSCC@epa.gov>  
**Sent:** Wednesday, November 20, 2024 10:33 AM  
**To:** Danganan, Eloise B <Eloise.Danganan@gdit.com>  
**Subject:** RE: REGION 5 | CASE 51834 | LAB ACE | SDG E2117 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

**This Message Is From an External Sender**

Please use caution with links, attachments, and any requests for credentials.

Hi Eloise,

Per the client, the lab's proposal is acceptable.

Thanks,  
Amanda

Amanda Wroble, Ph.D.  
Metals Chemist/Group Leader & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator  
USEPA Region 5 LSASD/AIMB  
Chicago Regional Laboratory  
536 S. Clark St. LAB-10C, 10<sup>th</sup> Floor  
Chicago, IL 60605  
(312) 353-0375

---

**From:** Danganan, Eloise B <[Eloise.Danganan@gdit.com](mailto:Eloise.Danganan@gdit.com)>  
**Sent:** Tuesday, November 19, 2024 2:23 PM  
**To:** R5RSCC <[R5RSCC@epa.gov](mailto:R5RSCC@epa.gov)>  
**Subject:** REGION 5 | CASE 51834 | LAB ACE | SDG E2117 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

**Caution:** This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon Region 5,

Please advise on the laboratory's reported issue for Case 51834/SDG E2117.

**ISSUE:** "Lab has received water samples for VOA analysis. Lab has performed the analysis for the samples E27M1, E27N7 & E27N8 in a continuous analytical sequence. Samples are found positive with high concentration of multiple analysis and required dilution analysis to bring target analytes within calibration range. In this case, instrument blank was analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted VOA analysis without instrument blank in between and further dilution in final electronic data.

Please see attachment for your reference."

Thank you,

ELOISE DANGANAN  
Environmental Analyst  
GDIT Federal Civilian Division  
Under contract to the EPA  
T 703.254.0067  
15036 Conference Center Drive  
Chantilly, VA 20151  
[www.gdit.com](http://www.gdit.com)

**GENERAL DYNAMICS**  
Information Technology

**LEAVE ALERT: 11/22-11/25, 11/28**

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

---

**From:** Sohil Jodhani <[Sohil.Jodhani@AllianceTG.com](mailto:Sohil.Jodhani@AllianceTG.com)>

**Sent:** Tuesday, November 19, 2024 1:20 PM

**To:** DASsupport <[DASsupport@gdit.com](mailto:DASsupport@gdit.com)>

**Cc:** Mohammad Ahmed <[mohammad.ahmed@alliancetg.com](mailto:mohammad.ahmed@alliancetg.com)>; Nimisha Pandya <[Nimisha.Pandya@AllianceTG.com](mailto:Nimisha.Pandya@AllianceTG.com)>

**Subject:** REGION 5 | CASE 51834 | LAB ACE | SDG E2117 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

**This Message Is From an External Sender**

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is sending this email with regards to case 51834 and SDG E2117.

Lab has received water samples for VOA analysis. Lab has performed the analysis for the samples E27M1, E27N7 & E27N8 in a continuous analytical sequence. Samples are found positive with high concentration of multiple analysis and required dilution analysis to bring target analytes within calibration range. In this case, instrument blank was analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted VOA analysis without instrument blank in between and further dilution in final electronic data.

Please see attachment for your reference.

**Thanks & Regards,**



**Sohil Jodhani**

**QA/QC Director**

**An Alliance Technical Group Company**

**Main:** 908-789-8900

**Direct:** 908-728-3152

**Address:** 284 Sheffield St, Ste 1, Mountainside, NJ 07092

[www.alliancetg.com](http://www.alliancetg.com)

